



CERTIFICATE OF CALIBRATION

Certificate No: 202311052

Instrument: EuroLab 240 Fully Automated Clinical Chemistry Analyzer

Customer Name: Bio Micro Laboratories , 67, Alagapuram Main Road, Opp Suramangalam
Police Station, Suramangalam, Salem - 636005.

Date of Installation: 14-08-2023

Serial No: AES2K3KT02182A

Alignment Check

Motion Check	Checked	Remark
Sample Probe	<input type="checkbox"/>	OK
Reagent Probe	<input type="checkbox"/>	OK
Mixing Bar	<input type="checkbox"/>	OK
Reaction Disk	<input type="checkbox"/>	OK
Reagent Disk	<input type="checkbox"/>	OK
Sample Disk	<input type="checkbox"/>	OK
Fluidic Disk	<input type="checkbox"/>	OK
Lamp	<input type="checkbox"/>	OK

System	Checked	Remark
Main unit	<input type="checkbox"/>	OK
Temp. unit	<input type="checkbox"/>	OK
Mixing unit	<input type="checkbox"/>	OK
Reaction unit	<input type="checkbox"/>	OK
Reagent unit	<input type="checkbox"/>	OK
Sample unit	<input type="checkbox"/>	OK



OPTICAL SECTION

Wavelength	Remark
340	OK
405	OK
450	OK
505	OK
546	OK
578	OK
630	OK
700	OK
Normal Range	20,000 to 62,000 Abs

Power Supply : 5V 12V 24V


The equipment is calibrated for the different parameters by using Spinreact Multi

Calibrator Name : **Spintrol H CAL** Lot.No:**2492** Exp: **2024-12**

This is to certify that the above mentioned product has been calibrated according to the standard operating procedure given by the **E-LAB Biological Science & Technology**.

Date of Calibration : **04 Nov 2023**

Next Due on: **03 Nov 2024**


Bharat Kumar
SSE

