

**MINDRAY BC-5130 CALIBRATION CERTIFICATE**

Customer Name	Biomicro Laboratories					
Department	Department of Hematology-Laboratory Services					
Address	No-67, Junction main road, Opp. Suramangalam police station, Suramangalam, Salem-636 005.					
	Date of Calibration	26-10-2023				
Instrument	BC-5130 Auto Haematology Analyzer	Next Calibration Date	25-10-2024			
Serial No	TR-81000491	Calibrator Lot Number	PLUS0923			
		Calibrator Expiry date	05-10-2023			
Calibration Values	Parameter	Target	Range	Mean of Measured Values	New Factor	Old Factor
	WBC x 10 <sup>9</sup> /L	9.00	8.80-9.20	8.97	97.84%	100%
	RBC x 10 <sup>12</sup> /L	4.60	4.52-4.68	4.64	100.77%	100%
	HGB g/dL	13.3	13.1-13.5	13.4	101.11%	100%
	MCV fl	86.7	84.7-88.7	86.3	99.59%	100%
	PLT x 10 <sup>9</sup> /L	253	241-265	256	101.56%	100%
This is to certify that the Mindray BC – 5130 Hematology Analyzer with Serial No: TR-81000491, Installed at Biomicro Laboratories has been calibrated using the standard reference material PLUS0923 Calibrator.						

Runned By :


*[Signature]*

Date: 01-11-2023

Application Specialist-IVD  
Golden Harvest Industries.

Reviewed By :

*[Signature]*



Date: 01-11-2023

Manager – Application  
Golden Harvest Industries.

**HEAD OFFICE:** First Floor, New No. 83, Old No. 33, New Boag Road, Opp. Kannammappet Corporation School, T.Nagar, Chennai - 600 017. INDIA Off: 044-24332994, Fax: 044-42868378.  
Www.goldenharvestindustries.com

**Toll Free Number: 1800 420 4444**