

CENTRIFUGE Calibration / Verification Record Sheet**Location: ICTC, CHC- KAMALAPURAM****Range : Up to 3500 RPM****Equipment ID:APSACS/ICTC/CHC/KMP/CEN-1****SL. No:2****Calibration Procedure Reference: In house verification of centrifuge using standard calibrated by Tachometer****Date of Calibration**

14-11-2023

Due Date of Calibration

13-11-2024

Certified Centrifuge ID: GMC-ATP/HIVTL/SRL/CEF/01**Tachometer ID: GMC-ATP/HIVTL/SRL/TM/01****Calibration date:**
19.05.2023**Due date of Calibration:**18.05.2024**Verification Done By: P. ANURADHA, Technical Officer, SRL, GMC, ANANTHAPURAMU****No.****STANDARD READING (RPM)****UUC READING (RPM)****ERROR (RPM)**

1.

499.3

500

0.7

2.

999.5

1000

0.5

3.

1998.2

2000

1.8

4.

2997.9

3000

2.1

5.

3496.3

3500

3.7

Accept/ Reject:Accepted since average difference is within +/-1000RPM**Action Taken: NA****Remarks: NA**

P. Anuradha
SRL-T.O
Dept. of Microbiology
GMC - Ananthapuram.