



CERTIFICATE OF CALIBRATION

FT-Q-25		Page 1 of 1	
ULR No. :	CC214423000000787F	Date of Issue :	10-02-2023
Certificate No. :	TVCSP/23/02/235-06	Recom. Due Date :	08-02-2024
Date of Calibration :	09-02-2023	Customer Details :	
M/s.FEROKE NEETHI MEDICAL LAB FEROKE KOZHICODE.		SRF No. :	235
		Calibrated at :	Lab
		Date of Receipt :	09-02-2023
		Cond. On Receipt :	Satisfactory
Details of Test Instrument:			
Description :	Refrigerator with Indicator	Model No. :	GL7292RPU/2018
Range :	2°C to 8°C	Serial No. :	--
Least Count :	0.1°C	Identification No. :	NML/EM/FA-7(11)
Make :	LG	Accuracy :	As Per Manual
Working range :	--	Location :	--
Details of Standard Used :			
Name	Certificate No.	Valid upto	Traceability
Standard RTD Sensor	CS/21/LB/TH/435-08	22-02-2023	R&D Instrument
Multi-Function Calibrator	TVCSP/22/0076/MFC-001	21-02-2023	ACVS
Work Instruction :	WI-T-02		
Environmental Details :	Temperature: 25±4 °C	Relative Humidity:	45-75 % RH

THERMAL CALIBRATION

(Temperature)

Calibration Results

Test Reading	Standard Reading Mean	Deviation	Expanded Uncertainty (±)	Coverage Factor
°C	°C	°C	°C	K
2	2.5	-0.5	2.5	2
4	4.4	-0.4	2.5	2
8	8.5	-0.5	2.5	2

Remarks:-

1. The above UUC was calibrated at Lab
2. UUC is defined as Unit Under Calibration
3. Reading given was the average of 3 Measurements
4. The measurement confidence level is 95.45%
5. The Calibration Certificate shall not be reproduced except in full without written approval of TVCSPL
6. The above Expanded Uncertainty includes Repeatability, Resolution & Uncertainty of Measurement
7. The above results relate only to the item calibrated
8. Temperature Scale : ITS-90


* End of Certificate*

Calibrated by :

K. Kameswaran
Mr.K.Kameswaran
(Calibration Engineer)

Authorised by:

Mr. Anand Manoharan
(QM & TM)



...redefining the true value

92, S.R.B Nagar Main Road,
Chennai - 600 099.
Tamil Nadu, India.

Ph: 044 - 4281 9208 / Cell: 94440 38069 /
97102 22422 / 97102 22522 / 97102 22622
Email : calibrationservices@live.com
www.truevaluecalibration.com
CIN No : U29268TN2015PTC103428



Quality is Assured