

VIEW LAB MONTHLY SUMMARY

Lab Name	RAJ CLINICAL LAB	Details About Robust Analysis									
Lab No	11847	Detail About Monthly Summary									
Month	October	Detail about SDI									
Year	2023										
Constituent Group	Chemistry II										
		Click on the analyte to view Graphical Data				All Analyser Result		Print			
		Date of Result Entered : 20/10/2023									
		Date of Report Published : 04/11/2023									
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	Participants SD	Your Value	SDI	U	
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	1673	129.03	9.72	12.54	93.8 mg/dL	-2.81	0.61	
2	UREA II	Others (Urease Berthelot / Ned Dye) II	Any Analyser (Automation / Semi Automation)	920	84.21	11.70	9.85	95.5 mg/dL	1.15	0.65	
3	CREATININE II	Jaffes End point II	Any Analyser (Automation / Semi Automation)	812	7.19	14.05	1.01	6.3 mg/dL	-0.88	0.07	
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Any Analyser (Automation / Semi Automation)	1325	8.35	11.84	0.99	7.2 mg/dL	-1.16	0.05	
5	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	1282	5.47	10.35	0.57	5.74 g/dL	0.48	0.03	
6	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	1253	3.29	9.69	0.32	3.9 g/dL	1.91	0.02	
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	1268	5.07	16.36	0.83	6.49 mg/dL	1.71	0.05	
8	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	1485	109.78	11.30	12.40	106 mg/dL	-0.30	0.64	
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	1243	104.67	16.11	16.86	116 mg/dL	0.67	0.96	
		SDI Range		Interpretation							
		Within -1.00 to +1.00		Excellent.							
		Within ±1.01 to ±2.00		Good.							
		Within ±2.01 to ±2.99		Accept with caution. Warning Signal.							
		Beyond ±3.0		Unacceptable performance. Action Signal.							

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out