

VIEW LAB MONTHLY SUMMARY

Lab Name	ASHOK COMPUTERISED CLINICAL LAB X-RAY AND ECG CENTRE		Details About Robust Analysis							
Lab No	11853		Detail About Monthly Summary							
Month	October		Detail about SDI							
Year	2023									
Constituent Group	Chemistry II									
Click on the analyte to view Graphical Data			All Analyser Result		Print					
			Date of Result Entered : 20/10/2023							
			Date of Report Published : 04/11/2023							
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	Participants SD	Your Value	SDI	U
1	GLUCOSE II	GOD-POD II	Biosystems	297	130.72	8.44	11.04	131 mg/dL	0.03	1.28
2	UREA II	Urease UV / GLDH II	Biosystems	236	86.66	10.28	8.91	95.4 mg/dL	0.98	1.16
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Biosystems	234	7.71	11.66	0.90	6.74 mg/dL	-1.08	0.12
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Biosystems	249	8.54	10.76	0.92	7.3 mg/dL	-1.35	0.12
5	T-PROTEIN II	Biuret - Colorimetric II	Biosystems	233	5.51	8.10	0.45	5.59 g/dL	0.18	0.06
6	ALBUMIN II	BCG - Colorimetric II	Biosystems	270	3.22	9.99	0.32	3.33 g/dL	0.34	0.04
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Biosystems	245	5.42	11.86	0.64	5.51 mg/dL	0.14	0.08
8	CHOLESTROL II	CHOD-PAP II	Biosystems	269	103.56	10.53	10.90	131.2 mg/dL	2.54	1.33
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Biosystems	230	108.12	12.08	13.06	129 mg/dL	1.60	1.72
SDI Range		Interpretation								
Within -1.00 to +1.00		Excellent.								
Within ±1.01 to ±2.00		Good.								
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

