

## CENTRIFUGE Calibration / Verification Record Sheet

Range : Up to 3500 RPM

Location: ICTC, CHC- LAKKIREDDIPALLI

SL. No:2

Equipment ID:APSACS/ICTC/CHC/LRPALLI/CEF/01

Calibration Procedure Reference: In house verification of centrifuge using standard calibrated by Tachometer

Date of Calibration

16-11-2023

Due Date of Calibration

15-11-2024

Certified Centrifuge ID: GMC-

Tachometer ID: GMC-  
ATP/HIVTL/SRL/TM/01Calibration date:  
19.05.2023Due date of  
Calibration:18.05.2024

ATP/HIVTL/SRL/CEF/01

ANANTHAPURAMU

Verification Done By: P. ANURADHA, Technical Officer, SRL, GMC, ANANTHAPURAMU

No.	STANDARD READING (RPM)	UUC READING (RPM)	ERROR (RPM)
1.	499.7	500	0.3
2.	999.2	1000	0.8
3.	1998.7	2000	1.3
4.	2997.5	3000	2.5
5.	3496.8	3500	3.2

Accept/ Reject: Accepted since average difference is within +/-1000RPM

Action Taken: NA

Remarks: NA

P. Anuradha  
SRL-7.0  
Dept. of Microbiology  
GMC - Ananthapuram