

CALIBRATION CERTIFICATE

This is to certify that Biochemistry Analyser SF-3000 PLUS bearing Serial No.4512-1154 which has been installed at SUGAM CLINICAL LAB & E.C.G CENTER, VILLUPURAM has been tested on the following aspects

1. LIGHT SOURCE :

- * Constant Light intensity for each filter.

2. CALIBRATION :

- * The Absorbance Factor according to the Factory set calibration.
- * The Read Cell Temperature according to the Factory set calibration.

3. FILTERS :

Quality of the individual filters viz.340, 405, 505, 545, 580 and 630 were checked for calibration through the instrument software and calibrated.

FILTERS(nm)	VOLTS(v)
340	2.075
405	2.702
505	3.245
545	3.412
580	3.426
630	3.543

On complete testing of the unit we found that the unit is working satisfactorily and used to carry out the necessary functions.

Date of Calibration : 21/02/2023

Next Calibration Due : 20/02/2024

For CPC DIAGNOSTICS PVT LTD.,


S. Manjula,
Deputy Manager,
Service Co-ordination.



CORPORATE OFFICE

No.70/6,Old No.108/6,4th floor,Westminister,Dr.Radhakrishnan Salai,Mylapore,Chennai-600004.
Phone No:044-23460168/169