

MK BEST CALIBRATION SERVICES



No. 27, F-2, 1st Floor, 2nd Street, Varalakshmi Nagar, Maduravoyal,
(Opp. MGR Engineering College), Chennai - 600 095.

Ph.: 044 - 23780211, Cell : 93802 66480 / 86958 18108 / 90032 77250

E-mail: mkbestcalibration@gmail.com, www.mkbestcalibrationservices.com

Certificate No. IQS/1710883



CERTIFICATE OF CALIBRATION

FF/7.8/01

Certificate No : MKBN/23/09/0406-001 Date of Issue : 23.09.2023
Date of Calibration : 21.09.2023 Due Date of Calibration : 20.09.2024

CUSTOMER INFORMATION

M/S. SUGAM CLINICAL LAB.,
NO.166/4, HOSPITAL ROAD,
VILLUPURAM.

UNIT UNDER CALIBRATION

Description	CENTRIFUGE
Range	3600 rpm
Resolution	100 rpm
Make	REMI
Model	C-854/4
Identification No	SCL/CB/GEN/01
Serial No	LCLC-3756
Location	LAB

S.No	Description	Id.No/SL No	Traceability	Certificate No	Validity
01	Digital Tachometer	MK/CAL-222	National Standards	CRMTL/01/422103978-A4	02.12.2023

ENVIRONMENTAL CONDITIONS

Temperature	24.2°C
Humidity	53 % RH

REFERENCE STANDARD

Uncertainty ± °C 1.57

CALIBRATION RESULTS

S.NO	Standard Reading rpm	UUC Readings rpm	Error rpm
1	509	500	-9
2	1013	1000	-13
3	2018	2000	-18
4	3023	3000	-23
5	3631	3600	-31

Remarks :

- The above UUC was Calibrated at Site.
- The Calibration Certificate Shall not be Reproduced Except In Full, Without Written Approval Of The Laboratory.
- The Recalibration Interval is Determined based on the User requirement.
- The Results Stated In This Certificate Relate Only to the Item Calibrated.
- The above Instrument used for Medical Purpose.

Calibrated by

S. Chandrabose

S.Chandrabose
(Calibration Engineer)

x-x-x-x End Of Certificate x-x-x-x



Authorised by

L. Magesh
L.Magesh
(MD/QM)