



# **CERTIFICATE OF CALIBRATION**

ADL/KO/SER/0262/23-24

Instrument : Mispa Clinia Fully Automated Biochemistry Analyser

**Customer Name** : Divine Touch Medi Clinic, Colonel Chowmuhani, Agartala, Tripura.

Date of Installation : 05/03/2021 S/N: **YM-84001332** 

# **Alignment Check**

Motion Check	Checked	Remark
Sample Probe		ОК
Reagent Probe		ОК
Mixer Probes		ОК
Reaction Disk		ОК
Reagent Disk		ОК
Sample Disk		ОК
Wash Station		ОК
Lamp		ОК

System	Checked	Remark
Main unit		ОК
Temp.unit		ОК
Mixing unit		OK
Reaction unit		OK
Reagent unit		ОК
Sample unit	П	OK

## AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN: U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India. Tel: + 91 484 286 7000 | Email: agappe@agappe.in

MUMBAI [REGISTERED OFFICE] 401 & 402, 4<sup>th</sup> Floor, Jaisingh Business Centre, 119, Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India. Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

**EQUIPMENT PLANT** X/588-CB, Block No. 32, KINFRA Small Industrial Park, Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

**DELHI OFFICE**DSM 540, 5<sup>th</sup> Floor, DLF Tower, Shivaji Marg,
New Delhi - 110 015, India.
Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

KOLKATA OFFICE 406, Merlin Matrix, Plot No-10, Block-DN, Sector V, Salt Lake City, Kolkata - 700 091. Tel: +91 33 4003 0451 | Email: kolkattaoffice@agappe.in

BANGALORE OFFICE Bandatone Ortice S-6, II\*S Floor, Red Cross Bhavan, No:26 Race Course Road, Bangalore - 560 001. Tel: +91 80 2228 8288 Email: bangaloreoffice@agappe.in





# **OPTICAL SECTION**

Photometer Unit Alignments	Status	Remarks
Lamp Brightness Adjustment		
Signal Collecting Position Adjustment		
Photoelectric Gain Adjustment		

The equipment is calibrated for the different parameters by using Multi Calibrator Lot.No:\_Exp:

This is to certify that the above-mentioned product has been calibrated according to the standard operating procedure given by the MINDRAY Medical Systems.

Date of Calibration: 12/SEP/2023

: 11/MAR/2024 **Next Due** 

For Agappe Diagnostics Ltd.



# **Authorised signatory**

Name: Debashis Bhattacharjee

Designation: Manager - Customer Service

AGAPPE DIAGNOSTICS LTD. 150 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India. Tel: + 91 484 286 7000 | Email: agappe@agappe.in

MUMBAI [REGISTERED OFFICE] 401 & 402, 4<sup>th</sup> Floor, Jaisingh Business Centre, 119, Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India. Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in **FOLLIPMENT PLANT** X/588-CB, Block No. 32, KINFRA Small Industrial Park Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

**DELHI OFFICE**DSM 540, 5<sup>th</sup> Floor, DLF Tower, Shivaji Marg,
New Delhi - 110 015, India.
Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

KOLKATA OFFICE 406, Merlin Matrix, Plot No-10, Block-DN, Sector V, Salt Lake City, Kolkata - 700 091. Tel: +91 33 4003 0451 | Email: kolkattaoffice@agappe.in

BANGALORE OFFICE BANGALORE OFFICE S-6, III\* Floor, Red Cross Bhavan, No:26 Race Course Road, Bangalore - 560 001. Tel: +91 80 2228 8288 Email: bangaloreoffice@agappe.in





# **INSTALLATION QUALIFICATION**

For

**FULLY AUTOMATED CLINICAL CHEMISTRY ANALYZER** 

**MODEL: MISPA CLINIA** 

AGAPPE DIAGNOSTICS LTD.

ADL/KO/SER/0085/23-24

ADL/FO/MCD/064/2015/R00

CORPORATE OFFICE / REAGENT PLANT Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India. Tel: + 91 484 286 7000 | Email: agappe@agappe.in

MUMBAI [REGISTERED OFFICE] 401 & 402, 4" Floor, Jaisingh Business Centre, 119, Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India. Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

FOLLIPMENT PLANT X/588-CB, Block No. 32, KINFRA Small Industrial Park Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN: U24239MH1998PLC115413

**DELHI OFFICE** DSM 540, 5th Floor, DLF Tower, Shivaji Marg, New Delhi - 110 015, India. Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in KOLKATA OFFICE 406, Merlin Matrix, Plot No-10, Block-DN, Sector V, Salt Lake City, Kolkata - 700 091. Tel: +91 33 4003 0451 | Email: kolkattaoffice@agappe.in

BANGALORE OFFICE S-6, II<sup>ee</sup> Floor, Red Cross Bhavan, No:26 Race Course Road, Bangalore - 560 001. Tel: +91 80 2228 8288 Email: bangaloreoffice@agappe.in





## **TABLE OF CONTENTS**

S No.	Contents	Page No.
I	Approval of the IQ procedure	3
II	Instructions	4
III	Scope	5
V	Installation Qualification	6
V	Comments	11
VI	System Certification	12

ADL/FO/MCD/064/2015/R00

KOLKATA OFFICE





# I. APPROVAL OF THE INSTALLATION QUALIFICATION PROCEDURE

Both **DIVINE TOUCH MEDI CLINIC, AGARTALA, TRIPURA** and **AGAPPE DIAGNOSTICS LTD**. are jointly responsible for the installation of **MISPA CLINIA** with serial number **YM-84001332** as per the attached Installation Qualification protocol.

Protocol Performed By: Representative of AGAPPE DIAGNOSTICS LTD.

Name	:	Debashis Bhattacharjee	Signatu	re:
Title	:	Manager - Customer Service	Date	:
Company	:	AGAPPE DIAGNOSTICS LTD.		



# Validation Team from DIVINE TOUCH MEDI CLINIC, AGARTALA, TRIPURA

Name	:	Signatur	e:
Designation	:	Date	:
Department	:		

## **Customer Authorizations:**

Name:

Title : INSTALLATION QUALIFICATION Signature :

Site : Date :

ADL/FO/MCD/064/2015/R00





# II. INSTRUCTIONS

- 1. An authorized AGAPPE DIAGNOSTICS LTD. representative will check and enter the specific data as outlined in the Installation Qualification. Each result will be noted and dated.
- 2. The concerned employees of **DIVINE TOUCH MEDI CLINIC, AGARTALA, TRIPURA** will verify each result and sign in each page. The member of the validation team will carry this out.
- 3. ALL deviations from the acceptance criteria detailed in this document will be noted in the COMMENTS section at the end of the IQ protocol. All resolution to such problems will also be noted in the COMMENTS section, and must be resolved prior to issuance of a SYSTEM CERTIFICATION.

ADL/FO/MCD/064/2015/R00

KOLKATA OFFICE

**FOUIPMENT PLANT** 





This Installation Qualification protocol will be performed on the MISPA CLINIA at DIVINE TOUCH MEDI CLINIC, AGARTALA, TRIPURA. This Installation protocol will define the documentation that will be used to evaluate the instruments installation in accordance with the manufacturer's specifications and intended use. Successful completion of this protocol will verify that the instrument identified has been installed in accordance with the intended usage.

Trained, knowledgeable personnel will perform qualification studies. Any exceptional conditions encountered during the qualification studies will be identified for review. Exceptional conditions will be investigated and the appropriate course of action determined. All documents will be initialed and dated.

ADL/FO/MCD/064/2015/R00

KOLKATA OFFICE





# SURGING IN OVATION STALLATION QUALIFICATION

A. <u>Instrument Identification</u> Verified Date: 05-03-2021

1. Model Name : MISPA CLINIA

2. Serial Number : YM-84001332

# B. Following is a list of accessories /consumables verified:

S No.	Component	Present (Yes/No)	Verified By	Comments
01	Analyzer Main unit			
02	Computer			
03	Printer			
04	RS232 Connector cable			
05	Power cord			
06	Waste can			
07	Water can			
08	Water connector tube			
09	Waste connector tube			
10	Water pre filter			
11	Probe assembly			
12	Mixer assembly			
13	Reagent tray			
14	Cuvette blocks			
15	Reagent bottles			

ADL/FO/MCD/064/2015/R00





# C. <u>Installation Checks</u>

Purpose: To make sure that the instrument is received with all the major components necessary for operation

S No.	Component	(Yes/No)	Verified By	Comments
1	Major components and accessories are present			
2	There is no physical damage to the components			
3	Ambient temperature is available			
4	Instrument is installed in the floor and leveled properly			
5	Power supply is correct			
6	Electric connections are tight, weatherproof & earthed			
7	Analyzer unit installed as per the manufacturers recommendations			
8	Water supply and waste drainage facilities checked			
9	Mispa Clinia Detergent Available			

## **Summary:**

The information is recorded against each data and the deviations if any is justified / explained properly.

Acceptance criteria:

**PARAMETER PASS FAIL** 

ADL/FO/MCD/064/2015/R00





# D. <u>Instrument Installation</u>

Date of Installation :05-03-2021

Installed by

S No.	Parameter	Done by	Comments
01	Installation of the main unit		
02	Installation of the computer system		
03	Installation of the printer		
04	Connecting the waste tubing		
05	Connecting the system water tubing		
06	Installing the reagent tray		
07	Installation of the probes		
08	Installation of the mixers		
09	Installing the cuvettes		
10	Installation of Wash Station		

Acceptance criteria: System should be "Ready" after daily maintenance without any error

**PASS FAIL PARAMETER** 

Parameter values for verification: System found "Ready" after daily maintenance

**FOUIPMENT PLANT** 

ADL/FO/MCD/064/2015/R00





Purpose: To check the cuvettes run Cuvette Check

**Procedure:** Manually put more than 40ml Mispa Clinia Detergent in position D of reagent disk and more than 5ml Mispa Clinia Detergent in position D2 of sample disk. Then select **Perform Cuvette Check** and click **Continue** to proceed.

## Acceptance criteria:

- No error Messages displayed
- All Cuvettes should show acceptance

PARAMETER PASS FAIL

Parameter values for verification: No Error Messages

ADL/FO/MCD/064/2015/R00

KOLKATA OFFICE

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN: U24239MH1998PLC115413

**FOUIPMENT PLANT** 





**Purpose:** To check the internal tubing's and flow of liquids

S No	Activity			Done By	y	Date
01	Do all checks i	n Hydro Unit				
gnment						
hotometric Un	it Reaction Unit	Sample Carousel	Rgt Carousel	Sample Probe Unit	Reagent Probe Unit	Sample Mixer Unit
Reagent Mixe Unit	Wash Station	Sample Barcode	Rgt Barcode	Hydro Unit	Empty Fluidic Tubes	Pyrology Unit
ISE Unit						
Ĩ	Supply/Drain Co	nformity Check				
2	Valve and Floate	Valve and Floater Test				
3	Pump Check	Pump Check				
4	Syringe Check					
5	Fluidic Prime	Fluidic Prime				
6	Water Tank Emp	Water Tank Empty Floater Check				
7	Wash Flow Test					
					Mech Reset	Exit

## Acceptance criteria:

- No error message or air bubbles in the internal tubing's
- All hydraulic units working well without air bubbles / leakage

PARAMETER PASS FAIL

Parameter values for verification: No Error Messages

ADL/FO/MCD/064/2015/R00

KOLKATA OFFICE

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN: U24239MH1998PLC115413

**FOUIPMENT PLANT** 

Email: bangaloreoffice@agappe.in

406, Merlin Matrix, Plot No-10, Block-DN, Sector V,

X/588-CB, Block No. 32, KINFRA Small Industrial Park





ADL/FO/MCD/064/2015/R00

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India. Tel: + 91 484 286 7000 | Email: agappe@agappe.in

MUMBAI [REGISTERED OFFICE]
401 & 402, 4° Floor, Jaisingh Business Centre, 119,
Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

FOLLIPMENT PLANT X/588-CB, Block No. 32, KINFRA Small Industrial Park, Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

**DELHI OFFICE** DSM 540, 5th Floor, DLF Tower, Shivaji Marg, New Delhi - 110 015, India. Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in KOLKATA OFFICE A06, Merlin Matrix, Plot No-10, Block-DN, Sector V, Salt Lake City, Kolkata - 700 091. Tel: +91 33 4003 0451 | Email: kolkattaoffice@agappe.in

BANGALORE OFFICE S-6, II<sup>nt</sup> Floor, Red Cross Bhavan, No:26 Race Course Road, Bangalore - 560 001. Tel: +91 80 2228 8288 Email: bangaloreoffice@agappe.in





# **VI. SYSTEM CERTIFICATION**

Study data has determined that the system described in this document either meets all criteria outline in this Installation Qualification Protocol, or exceptional conditions have been identified and documentation included. Exceptional conditions, if any, have been addressed. The system is ready for Operational Qualification.

Report Performed By: AGAPPE DIAGNOSTICS LTD. Representative

Name : Debashis Bhattacharjee

Designation : Manager - Customer Service

Signature :

Company : AGAPPE DIAGNOSTICS LTD. Date

Customer Authorizations: DIVINE TOUCH MEDI CLINIC, AGARTALA, TRIPURA

Name :

Designation : Signature :

Title : INSTALLATION QUALIFICATION Date :

ADL/FO/MCD/064/2015/R00





# **OPERATIONAL QUALIFICATION**

For

**FULLY AUTOMATED CLINICAL CHEMISTRY ANALYZER** 

**MODEL: MISPA CLINIA** 

AGAPPE DIAGNOSTICS LTD.

ADL/KO/SER/0086/23-24

ADL/FO/MCD/065/2015/R00

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN: U24239MH1998PLC115413

**CORPORATE OFFICE / REAGENT PLANT** Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India. Tel: + 91 484 286 7000 | Email: agappe@agappe.in

401 & 402, 4th Floor, Jaisingh Business Centre, 119, Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.

Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

MUMBAI [REGISTERED OFFICE]

X/588-CB, Block No. 32, KINFRA Small Industrial Park, Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

DSM 540, 5th Floor, DLF Tower, Shivaji Marg, New Delhi - 110 015, India. Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

406, Merlin Matrix, Plot No-10, Block-DN, Sector V, Salt Lake City, Kolkata - 700 091. Tel: +91 33 4003 0451 | Email: kolkattaoffice@agappe.in





# **TABLE OF CONTENTS**

S No.	Contents	Page No.
I	Approval of the OQ procedure	3
II	Instructions	4
III	Scope	5
IV	Operational Qualification	6
V	Comments	13
VI	System Certification	14

ADL/FO/MCD/065/2015/R00

401 & 402, 4" Floor, Jaisingh Business Centre, 119, Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India. Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

MUMBAI [REGISTERED OFFICE]





# I. APPROVAL OF THE OPERATIONAL QUALIFICATION PROCEDURE

Both DIVINE TOUCH MEDI CLINIC, AGARTALA, TRIPURA and AGAPPE DIAGNOSTICS LTD. are jointly responsible for the installation of MISPA CLINIA with serial number YM-84001332 as per the attached OPERATIONAL QUALIFICATION protocol.

Protocol Performed By: Representative of AGAPPE DIAGNOSTICS LTD.

Name	:	Debashis Bhattacharjee	Signature	:
Designation	:	Manager - Customer Service	Date	:
Company	:	AGAPPE DIAGNOSTICS LTD.		



## Validation Team from DIVINE TOUCH MEDI CLINIC, AGARTALA, TRIPURA

Name	:	Signatur	e :
Designation	:	Date	:
Department	:		

## **Customer Authorizations:**

Name:

Title : OPERATIONAL QUALIFICATION Signature

Site Date

ADL/FO/MCD/065/2015/R00

406, Merlin Matrix, Plot No-10, Block-DN, Sector V,

AGAPPE DIAGNOSTICS LTD. 150 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

Email: bangaloreoffice@agappe.in





# II. INSTRUCTIONS

- An authorized AGAPPE DIAGNOSTICS LTD representative will check and enter the specific data as outlined in the OPERATIONAL QUALIFICATION. Each result will be noted and dated.
- 2. ALL deviations from the acceptance criteria detailed in this document will be noted in the COMMENTS section at the end of the OQ protocol. All resolution to such problems will also be noted in the COMMENTS section, and must be resolved prior to issuance of a SYSTEM CERTIFICATION.

ADL/FO/MCD/065/2015/R00





This Operational Qualification protocol will be performed on the MISPA CLINIA at DIVINE TOUCH MEDI CLINIC, AGARTALA, TRIPURA. This Operational protocol will define the documentation that will be used to evaluate the instruments installation in accordance with the manufacturer's specifications and intended use. Successful completion of this protocol will verify that the instrument identified has been installed in accordance with the intended usage. Trained, knowledgeable personnel will perform qualification studies. Any exceptional conditions encountered during the qualification studies will be identified for review. Exceptional conditions will be investigated and the appropriate course of action determined. All documents will be initialed and dated.

ADL/FO/MCD/065/2015/R00





# IV. OPERATIONAL QUALIFICATION

A. Instrument Identification

Verified Date:

1. Model Name : MISPA CLINIA

2. Serial Number : YM-84001332

B. Following is a list of tests to be performed and verified:

Test No.	Test Name	Test purpose	Verified By and date
01	Start up	To make the equipment ready for operation	
02	Daily maintenance	To clean appropriate modules	
03	Calibration	To calibrate the system	
04	QC check	To confirm that the system is calibrated and working within specifications	
05	Reproducibility check	To check the precision [CV %] after calibration	
06	Sample programming and Analysis	To run the samples	
07	Shut down procedure	To shut down the system	

ADL/FO/MCD/065/2015/R00





# **Test 01: Starting the system**

**Purpose:** To make the instrument READY for operation

## **Summary:**

Instrument checks functioning of different parts of the instrument automatically. If there is an error code, initialize the system and follow corrective action instructions provided for the error message.

#### Procedure:

- Wait for the instrument to get ready after initialization
- Check the room temperature and switch on the Air Conditioner
- Check the UPS.
- Switch on the MISPA CLINIA by pressing the main switch, then switch on the computer and then the system power
- Double click the analyzer icon to initialize the system software
- Input the password and do as prompted, the system menu pops up if start up is finished
- If not, initialize by again after solving the error displayed
- Follow instructions provided for the error message

Acceptance criteria: System to display READY status

**PARAMETER PASS FAIL** 

Parameter values for verification: "READY" on Status Area

ADL/FO/MCD/065/2015/R00

X/588-CB. Block No. 32. KINERA Small Industrial Park

Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.





# **Test 02: Daily Maintenance**

Purpose: To clean appropriate modules so as per the daily maintenance protocol

Method:

Refer detailed procedure for Daily Maintenance in the software

S No.	Activity	Done by	Date
01	Empty waste container		
02	Check Wash solutions		
03	Check system water		
04	Check Reagents		
05	Clean Reagent Block		
06	Clean out side area		
07	Load supplies and remove outdated and empty reagents		
08	Check main menu screen		
09	Perform Quality Control		_

Acceptance criteria: System should be "Ready" after daily maintenance without any error

**PASS FAIL PARAMETER** 

Parameter values for verification: System found "Ready" after daily maintenance

ADL/FO/MCD/065/2015/R00

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN: U24239MH1998PLC115413

Email: bangaloreoffice@agappe.in

S-6, II<sup>nd</sup> Floor, Red Cross Bhavan, No:26 Race Course Road, Bangalore - 560 001. Tel: +91 80 2228 8288





Test 03: Calibration of the system

**Purpose:** To calibrate the system

Procedure:

S No	Activity	Done By	Date
01	Preparation of the cal material		
02	Performing Calibration with calibration programming screen		

Acceptance criteria: Calibration data shows concordance OD values in triplicate measurement

**PARAMETER PASS FAIL** 

Parameter values for verification : No error message in the printout

Test 04: QC check

Purpose: To confirm that systems, reagents and consumables are acceptable & working within specifications for each assay used.

## **Procedure:**

S No.	Activity	Done By	Date
01	Preparing Biochemistry control material		
02	Creating QC file		
03	QC sample programming and analysis		

Acceptance criteria: QC results within specified limits mentioned on the control product insert

**PARAMETER PASS FAIL** 

Parameter values for verification: QC values within ± 2SD

ADL/FO/MCD/065/2015/R00





Test 05: Reproducibility Check

**Purpose:** To check the reproducibility of instrument

S No.	Activity	Done By	Date
01	Preparing control material/ sample		
02	Running triplicates		

Acceptance criteria: Results CV within specified limits

**PARAMETER PASS FAIL** 

Parameter values for verification: CV% within the limits

**Test 06: Sample programming and Analysis** 

**Purpose:** To run the samples

Procedure:

S No.	Activity	Done By	Date
01	Preparing and Processing of samples		
02	Programming samples		
03	Aspirating the samples		
04	Viewing samples in process		
05	Review results: Monitoring results		

Acceptance criteria: Samples Analysis without any error

**PARAMETER PASS FAIL** 

Parameter values for verification: Sample analysis without any error

ADL/FO/MCD/065/2015/R00

X/588-CB, Block No. 32, KINFRA Small Industrial Park

Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.





Purpose: To shut down the system

**Procedure:** 

S No.	Activity	Done By	Date
01	Preparing the system for shut down		
02	Following the on screen instructions		
03	Switch off the instrument when prompted		

Acceptance criteria: Shut down without any error

**PARAMETER PASS FAIL** 

Parameter values for verification: Shut down without any error

# V. OPERATIONAL PROCEDURE

# **Certificate of Training**

## **Technician Training**

This certifies that the technicians have received basic user training in the following categories for the system described in this Operational Qualification.

S No.	Training program	Initials	Date
1	Instrument Setup		
2	System Operation		
3	Basic trouble shooting and Maintenance		

ADL/FO/MCD/065/2015/R00





# 2. Operator Training

The users responsible for the operation of this equipment have been trained in the proper usage of the system. Training focused on the basic operation and maintenance of the system.

S No.	Operators	Department	Initials	Date

ADL/FO/MCD/065/2015/R00

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

401 & 402, 4th Floor, Jaisingh Business Centre, 119, Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India. Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

MUMBAI [REGISTERED OFFICE]





# **VI. COMMENTS**

ADL/FO/MCD/065/2015/R00

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India. Tel: + 91 484 286 7000 | Email: agappe@agappe.in

MUMBAI [REGISTERED OFFICE] 401 & 402, 4" Floor, Jaisingh Business Centre, 119, Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India. Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in X/588-CB, Block No. 32, KINFRA Small Industrial Park, Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

DELHI OFFICE DSM 540, 5<sup>th</sup> Floor, DLF Tower, Shivaji Marg, New Delhi - 110 015, India. Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

A06, Merlin Matrix, Plot No-10, Block-DN, Sector V, Salt Lake City, Kolkata - 700 091. Tel: +91 33 4003 0451 | Email: kolkattaoffice@agappe.in

BANGALURE OFFICE 5-6, II<sup>as</sup> Floor, Red Cross Bhavan, No:26 Race Course Road, Bangalore - 560 001. Tel: +91 80 2228 8288 Email: bangaloreoffice@agappe.in





# VII. SYSTEM CERTIFICATION

Study data has determined that the system described in this document either meets all criteria outline in this Operational Qualification Protocol, or exceptional conditions have been identified and documentation included. Exceptional conditions, if any, have been addressed. The system is ready for Performance Qualification.

Report Performed By: AGAPPE DIAGNOSTICS LTD. Representative

Name : Debashis Bhattacharjee

Designation : Manager - Customer Service

Signature

Company : AGAPPE DIAGNOSTICS LTD. Date

Customer Authorizations: DIVINE TOUCH MEDI CLINIC, AGARTALA, TRIPURA

Name

Designation Signature

Title : OPERATIONAL QUALIFICATION Date

ADL/FO/MCD/065/2015/R00

AGAPPE DIAGNOSTICS LTD. 150 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

MUMBAI [REGISTERED OFFICE]

Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.ir

Salt Lake City, Kolkata - 700 091. Tel: +91 33 4003 0451 | Email: kolkattaoffice@agappe.in

406, Merlin Matrix, Plot No-10, Block-DN, Sector V,





ADL/FO/MCD/065/2015/R00

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN: U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India. Tel: + 91 484 286 7000 | Email: agappe@agappe.in

MUMBAI [REGISTERED OFFICE] 401 & 402, 4" Floor, Jaisingh Business Centre, 119, Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India. Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in X/588-CB, Block No. 32, KINFRA Small Industrial Park, Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

DELHI OFFICE DSM 540, 5<sup>th</sup> Floor, DLF Tower, Shivaji Marg, New Delhi - 110 015, India. Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

A06, Merlin Matrix, Plot No-10, Block-DN, Sector V, Salt Lake City, Kolkata - 700 091. Tel: +91 33 4003 0451 | Email: kolkattaoffice@agappe.in

BANGALURE OFFICE 5-6, II<sup>as</sup> Floor, Red Cross Bhavan, No:26 Race Course Road, Bangalore - 560 001. Tel: +91 80 2228 8288 Email: bangaloreoffice@agappe.in





# PERFORMANCE QUALIFICATION

For

**FULLY AUTOMATED CLINICAL CHEMISTRY ANALYZER** 

**MODEL: MISPA CLINIA** 

AGAPPE DIAGNOSTICS LTD.

ADL/KO/SER/0087/23-24

ADL/FO/MCD/066/2015/R00

AGAPPE DIAGNOSTICS LTD. 150 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

**KOLKATA OFFICE** 

CORPORATE OFFICE / REAGENT PLANT Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India. Tel: + 91 484 286 7000 | Email: agappe@agappe.in

MUMBAI [REGISTERED OFFICE]

Mumbai (REUSTERE OFFICE) 401 & 402, 4" Floor, Jaisingh Business Centre, 119, Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India. Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

X/588-CB, Block No. 32, KINFRA Small Industrial Park, Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

**EQUIPMENT PLANT** 

406, Merlin Matrix, Plot No-10, Block-DN, Sector V, Salt Lake City, Kolkata - 700 091. Tel: +91 33 4003 0451 | Email: kolkattaoffice@agappe.in BANGALORE OFFICE





# **TABLE OF CONTENTS**

S No.	Contents	Page No
I	Approval of the PQ procedure	3
II	Instructions	4
III	Scope	5
IV	Performance Qualification	6
V	Comments	15
VI	System Certification	16

ADL/FO/MCD/066/2015/R00

DELHI OFFICE





# I. APPROVAL OF THE PERFORMANCE QUALIFICATION PROCEDURE

Both DIVINE TOUCH MEDI CLINIC, AGARTALA, TRIPURA and AGAPPE DIAGNOSTICS LTD. are jointly responsible for the installation of MISPA CLINIA with serial number YM-84001332 as per the attached PERFORMANCE Qualification protocol.

Protocol Performed By: Representative of AGAPPE DIAGNOSTICS LTD.

Name	:	Debashis Bhattacharjee	Signature:
Designation	:	Manager - Customer Service	Date :
Company	:	AGAPPE DIAGNOSTICS LTD.	



## Validation Team from DIVINE TOUCH MEDI CLINIC, AGARTALA, TRIPURA

Name	:	Signature :	
Designation	:	Date	:
Department	:		

## **Customer Authorizations:**

Name:

Title : PERFORMANCE QUALIFICATION Signature

Site Date

ADL/FO/MCD/066/2015/R00





# II. INSTRUCTIONS

- 1. An authorized AGAPPE DIAGNOSTICS LTD. representative will check and enter the specific data as outlined in the Performance Qualification. Each result will be noted and dated.
- 2. The concerned employees of **DIVINE TOUCH MEDI CLINIC, AGARTALA, TRIPURA** will verify each result and sign in each page. The member of the validation team will carry this out.
- 3. All deviations from the acceptance criteria detailed in this document will be noted in the COMMENTS section at the end of the PQ protocol. All resolution to such problems will also be noted in the COMMENTS section, and must be resolved prior to issuance of a SYSTEM CERTIFICATION.

ADL/FO/MCD/066/2015/R00

**KOLKATA OFFICE** 





This Performance Qualification protocol will be performed on the MISPA CLINIA at DIVINE TOUCH MEDI CLINIC, AGARTALA, TRIPURA. This Performance protocol will define the documentation that will be used to evaluate the instruments installation in accordance with the manufacturer's specifications and intended use. Successful completion of this protocol will verify that the instrument identified has been installed in accordance with the intended usage. Trained, knowledgeable personnel will perform qualification studies. Any exceptional conditions encountered during the qualification studies will be identified for review. Exceptional conditions will be investigated and the appropriate course of action determined. All documents will be initialed and dated.

ADL/FO/MCD/066/2015/R00

Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.

MIIMBAI [REGISTERED OFFICE]

401 & 402, 4th Floor, Jaisingh Business Centre, 119,

Tel: +91 22 4300 8000 | Email: mumbaioffice@aga





# IV. PERFORMANCE QUALIFICATION

A. <u>Instrument Identification</u> Verified Date:

> 1. Model Name : MISPA CLINIA 2. Serial Number : YM-84001332

## B. Following is a list of tests to be performed and verified:

1	Test No.	Test Name	Test purpose	Verified By and date
	01	Start up	To make the equipment ready for operation	
\	02	Daily maintenance	To clean appropriate modules	
	03	Cell blank	To check the cuvettes	
	04	Priming	To check the tubing's, and flow	
	05	Auto gain	To check the cuvette and photometry system	
	06	Calibration	To calibrate the system	
	07	QC check	To confirm that the system and reagents are acceptable and working within specifications	
	08	Reproducibility check	To check the precision [CV %]	
	09	Sample programming and Analysis	To run the samples	
	10	Shut down procedure	To shut down the system	

ADL/FO/MCD/066/2015/R00





# Test 01: Starting the system

**Purpose:** To make the instrument READY for operation

#### **Summary:**

Instrument checks functioning of different parts of the instrument automatically; if there is an error code, initialize the system and follow corrective action instructions provided for the error message.

#### **Procedure:**

- Wait for the instrument to get ready after initialization
- Check the room temperature and switch on the Air Conditioner
- Check the UPS.
- Switch on the MISPA CLINIA by pressing the main switch, then switch on the computer and then the system power
- Double click the analyzer icon to initialize the system software
- Input the password and do as prompted, the system menu pops up if start up is finished
- If not, initialize by again after solving the error displayed
- Follow instructions provided for the error message

Acceptance criteria: System to display READY status

**PARAMETER PASS FAIL** 

Parameter values for verification: "READY" on Status Area

ADL/FO/MCD/066/2015/R00





# **Test 02: Daily Maintenance**

Purpose: To clean appropriate modules so as per the daily maintenance protocol

Method:

Refer detailed procedure for Daily Maintenance

S No.	Activity	Done by	Date
01	Empty waste container		
02	Check Wash solutions		
03	Check system water		
04	Check Reagents		
05	Clean Reagent Block		
06	Clean out side area		
07	Load supplies and remove outdated and empty reagents		
08	Check main menu screen		
09	Perform Quality Control		

Acceptance criteria: System should be "Ready" after daily maintenance without any error

**PARAMETER PASS FAIL** 

Parameter values for verification: System found "Ready" after daily maintenance

ADL/FO/MCD/066/2015/R00





# **Test 03: Photometer Check**

**Purpose:** To check the photometer after running some cycles of tests

#### **Photometer Check**

- 1. Execute Photometer Check.
- 2. Make sure that the lamp has been turned on for over 10 minutes. Operate according to screen prompts.
- 3. If the Status field shows Normal, it indicates that the lamp's light intensity satisfies the requirements of measurement; if it shows "Light intensity is weak" in red, it indicates that the lamp has insufficient light intensity.
- 4. If an alarm occurs during the check, operate as follows:
  - If an alarm indicates the lamp is off, check if the lamp has been turned on.
  - If the alarm indicates light intensity too strong
  - If the alarm indicates light intensity weak, replace the lamp immediately.

## Acceptance criteria:

- No error Messages displayed
- All checks acceptance

**PARAMETER PASS** FAIL

Parameter values for verification: No Error Messages

ADL/FO/MCD/066/2015/R00





**Test 04: Priming** 

**Purpose:** To check the internal tubing's and flow of liquids

S No	Activity		Don	e By	Date
01	Do all checks in Hydro Unit				
nment					
notometric Unit	Reaction Unit Sample Carousel Rgt Carousel	Sample Probe Unit	Reagent Probe Unit	Sample Mixer Unit	
Reagent Mixer Unit	Wash Station Sample Barcode Rgt Barcode	Hydro Unit	Empty Fluidic Tubes	Pyrology Unit	
ISE Unit					
1	Supply/Drain Conformity Check				
2	Valve and Floater Test	Valve and Floater Test			
3	Pump Check				
4	Syringe Check				
5	Fluidic Prime				

## Acceptance criteria:

- No error message or air bubbles in the internal tubing's
- All hydraulic units working well without air bubbles / leakage

PARAMETER PASS **FAIL** 

Water Tank Empty Floater Check

Wash Flow Test

Parameter values for verification: No Error Messages

ADL/FO/MCD/066/2015/R00

AGAPPE DIAGNOSTICS LTD. 150 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413


Mech Reset

Exit





#### **Test 05: Workflow**



Acceptance criteria: The workflow should be correct in order

**PARAMETER** 

**PASS** 

**FAIL** 

Parameter values for verification

: No unwanted stopping in between

ADL/FO/MCD/066/2015/R00





Test 07: QC check

**Purpose:** To confirm that systems, reagents and consumables are acceptable & working within specifications for each assay used.

#### **Procedure:**

S No.	S No. Activity		Date
01	Preparing Biochemistry control material		
02	Creating QC file		
03	QC sample programming and analysis		

Acceptance criteria: QC results within specified limits mentioned on the control product insert

**PARAMETER PASS FAIL** 

Parameter values for verification: QC values within ± 2SD

**Test 08: Reproducibility Check** 

Purpose: To check the reproducibility of instrument

S No.	Activity	Done By	Date
01	Preparing control material/ sample		
02	Running triplicates		

Acceptance criteria: Results CV within specified limits

**PARAMETER PASS FAIL** 

**Parameter values for verification:** CV% within the limits

ADL/FO/MCD/066/2015/R00





**Purpose:** To run the samples

**Procedure:** 

S No.	Activity	Done By	Date
01	Preparing and Processing of samples		
02	Programming samples		
03	Aspirating the samples		
04	Viewing samples in process		
05	Review results: Monitoring results		

Acceptance criteria: Samples Analysis without any error

**PARAMETER PASS FAIL** 

Sample analysis without any error Parameter values for verification:

Test 10: Shut down

Purpose: To shut down the system

**Procedure:** 

5	S No.	Activity	Done By	Date
	01	Preparing the system for shut down		
	02	Following the on screen instructions		
	03	Switch off the instrument when prompted		

Acceptance criteria: Shut down without any error

**PARAMETER PASS FAIL** 

**Parameter values for verification:** Shut down without any error.

ADL/FO/MCD/066/2015/R00





# **OPERATIONAL PROCEDURE**

# **Certificate of Training**

## 1. Technician Training

This certifies that the technicians have received basic user training in the following categories for the system described in this Operational Qualification.

S No.	Training program	Initials	Date
1.	Instrument Setup		
2.	System Operation		
3.	Basic trouble shooting and Maintenance		

# 2. Operator Training

The users responsible for the operation of this equipment have been trained in the proper usage of the system. Training focused on the basic operation and maintenance of the system.

S No.	Operators	Department	Initials	Date

ADL/FO/MCD/066/2015/R00





ADL/FO/MCD/066/2015/R00

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

**CORPORATE OFFICE / REAGENT PLANT** Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India. Tel: + 91 484 286 7000 | Email: agappe@agappe.in

MUMBAI [REGISTERED OFFICE] Mumbai (REUSTERE OFFICE) 401 & 402, 4" Floor, Jaisingh Business Centre, 119, Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India. Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in **EQUIPMENT PLANT** X/588-CB, Block No. 32, KINFRA Small Industrial Park, Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

DELHI OFFICE DSM 540, 5th Floor, DLF Tower, Shivaji Marg, New Delhi - 110 015, India. Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

KOLKATA OFFICE 406, Merlin Matrix, Plot No-10, Block-DN, Sector V, Salt Lake City, Kolkata - 700 091. Tel: +91 33 4003 0451 | Email: kolkattaoffice@agappe.in

BANGALORE OFFICE S-6, II<sup>nd</sup> Floor, Red Cross Bhavan, No:26 Race Course Road, Bangalore - 560 001. Tel: +91 80 2228 8288 Email: bangaloreoffice@agappe.in





# VI. SYSTEM CERTIFICATION

Study data has determined that the system described in this document either meets all criteria outline in this Performance Qualification Protocol, or exceptional conditions have been identified and documentation included. Exceptional conditions, if any, have been addressed. The system is ready for specified usage.

Report Performed By: AGAPPE DIAGNOSTICS LTD. Representative

Name : Debashis Bhattacharjee

Designation : Manager - Customer Service Signature

Company : AGAPPE DIAGNOSTICS LTD. Date

Customer Authorizations: DIVINE TOUCH MEDI CLINIC, AGARTALA, TRIPURA

Name

Designation Signature

Title : PERFORMANCE QUALIFICATION Date

ADL/FO/MCD/066/2015/R00