



CALIBRATION REPORT

AGAPPE

ADL/MO/SER/0230/23-24

Customer: UPC Diagnostic Centre, Vadamadurai.

Serial No. 212030336568

Model : Mispa Viva

DOI : 23/04/2021

Date of Calibration : 03/11/2023

INSTRUMENT CONTROL

Visual Control: PASSED

Zero Drift after 1 Hour of zero setting (WL=340nm) 0.001

PHOTOMETRIC LINEARITY

Solution	WL (nm)	100%	75%	50%	25%
Potassium dichromate	340nm	100.2	74.6	51.8	25.2
P-Nitrophenol	405nm	100.4	74.5	50.3	25.8
Cobaltous Ammonium Sulphate	505nm	100.8	76.9	50.1	24.7

CHEMISTRY CONTROL

	TEST	REAGENT	CONTROL SERUM	THEORETICAL VALUE	READ VALUE	RANGE
END POINT	Triglyceride	Triglyceride	Lot No: 32307084 Expiry: 02/2025	435	398	358 - 512
KINETIC	Glucose	Glucose	Lot No: 32307084 Expiry: 02/2025	214	244.25	169 - 259
FIXED TIME	Creatinine	Creatinine	Lot No: 32307084 Expiry: 02/2025	2.35	2.25	1.75 - 2.95

Validity: 6 months from the date of calibration – 02/05/2024

For, Agappe Diagnostics Ltd,
Authorized Signatory



Name: Anbarasu.R

Designation: Assistant Manager - Customer service

AGAPPE DIAGNOSTICS LTD.

ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT

Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
Tel: +91 484 286 7000 | Email: agappe@agappe.in

EQUIPMENT PLANT

X/588-CB, Block No. 32, KINFRA Small Industrial Park,
Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

KOLKATA OFFICE

406, Merlin Matrix, Plot No-10, Block-DN, Sector V,
Salt Lake City, Kolkata - 700 091.
Tel: +91 33 4003 0451 | Email: kolkatoffice@agappe.in

MUMBAI (REGISTERED OFFICE)

401 & 402, 4th Floor, Lakshmi Business Centre, 119

DELHI OFFICE

DSM 540, 5th Floor, DLF Tower, Shivaji Marg.

BANGALORE OFFICE

5-6, 11th Floor, Red Cross Bhavan, No-26 Race Course Road.