



Ref No. TBMSZ1/SSL/10005
Date : 01. 03 2023

PJ Clinical Laboratory,
No 3 Poonthottam Pathai,
Villupuram

Dear Sirs,

As desired by you please find the Calibration Check Report on Chem-7
Sr No. N120118

1. Year of manufacturing : 2017
2. Range of the instrument: Optical density Linear form 0-2.5
3. Date of Calibration : 01. 03 2023
4. Next Calibration Due: 28.02.2024

Test Voltages / Test Temperature -

| Test Voltage | Actual Voltage | Ok / Not Ok |
|---------------------|----------------|-------------|
| Logic 5 V | 5.10V | OK |
| Line voltage | 220V | OK |
| Lamp Voltage 12v | 11.7V | OK |
| Temperature 25°C | 25°C | OK |
| Temperature 30°C | 30°C | OK |
| Temperature 37°C | 37°C | OK |
| Input Voltage | 18.0 VDC | OK |
| Pump Calibration | 8541 | OK |



UNMATCHED SERVICE
SINCE 1979.

| Sr.No. | Filters | Pre Channel | Log Channel |
|--------|---------|-------------|-------------|
| 1 | 340 | 40k | 0.77 |
| 2 | 405 | 39k | 0.71 |
| 3 | 450 | 43k | 0.67 |
| 4 | 505 | 62k | 0.51 |
| 5 | 546 | 38k | 0.72 |
| 6 | 578 | 28k | 0.85 |
| 7 | 630 | 68k | 0.42 |
| 8 | 670 | 88k | 0.33 |

Thanking you and assuring of our prompt attention at all times.

Yours Sincerely,
For **TRANSASIA BIO-MEDICALS LTD.,**

S.SIVA PANDI
SERVICE ENGINEER