

MK BEST CALIBRATION SERVICES



No. 27, F-2, 1st Floor, 2nd Street, Varalakshmi Nagar, Maduravoyal,
(Opp. MGR Engineering College), Chennai - 600 095.

Ph.: 044 - 23780211, Cell : 93802 66480 / 86958 18108 / 90032 77250

E-mail: mkbestcalibration@gmail.com, www.mkbestcalibrationservices.com



CERTIFICATE OF CALIBRATION



FF/7.8/01

Certificate No	: MKBN/23/09/0421-001	Date of Issue	: 25.09.2023
Date of Calibration	: 22.09.2023	Due Date of Calibration	: 21.09.2024

CUSTOMER INFORMATION		UNIT UNDER CALIBRATION	
M/S. P J CLINICAL LABORATORY., NO.3, POONTHOTTAM PATHAL, VILLUPURAM - 605602.		Description	CENTRIFUGE
		Range	3600 rpm
		Resolution	100 rpm
		Make	REMI
		Model	C 852
		Identification No	PJL/CB/GEN/01
		Serial No	KJLC-7578
		Location	LAB

S.No	Description	Id.No/SL No	Traceability	Certificate No	Validity
01	Digital Tachometer	MK/CAL-222	National Standards	CRMTL/01/422103978-A4	02.12.2023

ENVIRONMENTAL CONDITIONS		REFERENCE STANDARD	
Temperature	23.6°C	Uncertainty ± °C	1.57
Humidity	54 % RH		

CALIBRATION RESULTS

S.NO	Standard Reading rpm	UUC Readings rpm	Error rpm
1	505	500	-5
2	1017	1000	-17
3	2025	2000	-25
4	3031	3000	-31
5	3639	3600	-39

Remarks :

- The above UUC was Calibrated at Site.
- The Calibration Certificate Shall not be Reproduced Except In Full, Without Written Approval Of The Laboratory.
- The Recalibration Interval is Determined based on the User requirement.
- The Results Stated In This Certificate Relate Only to the Item Calibrated.
- The above Instrument used for Medical Purpose.

Calibrated by

S.Chandrabose
(Calibration Engineer)

X-X-X-X- End Of Certificate -X-X-X-X



Authorised by

L.Magesh
(MD/QM)