




CALIBRATION CERTIFICATE

Page 1 of 1

1. Customer Name:	M/s. CHRONEVA HEALTH LAB PRIVATE LIMITED, The Orclid No-68, Ground Floor, 9th main, HMT Layout, RT Nagar, Bengaluru-560032.	6. Certificate No:	RCPL/23-24/ML/I5294-02				
		7. ULR No:	CC221523000014453F				
		8. Date of Issue:	02-11-2023				
2. SRF No:	I5294	9. Date of Calibration:	26-10-2023				
3. Date of SRF/Receipt:	18-10-2023	10. Next Cal Due:	25-10-2024				
4. Discipline:	Mechanical Calibration	11. Calibration SOP no:	RCPL/SOP/ML-21				
5. Calibrated At:	Mass & Volume Lab	12. Condition of DUC on Receipt:	Satisfactory				
13. ENVIRONMENTAL CONDITION : Temperature: 20.84 °C , Humidity: 58.33 % RH Ambient Pressure : 911.5 hpa, Water Temperature: 20.2 °C							
14. DUC DETAILS:							
Nomenclature :	Micropipette	Serial No:	SW13151				
Make & Model :	Thermo Scientific & Finnppipette F3	Code/ID No :	CHVL_BLR_MP1_006				
Range :	5 - 50 µl	Resolution:	0.1 µl				
15. STANDARD INSTRUMENT USED:							
Nomenclature:	Electronic Semi Micro Balance	Serial No:	33604236				
Make / Model :	Sartorius / CPA225D	Cal Due Date:	10-09-2024				
Range :	0 to 220g	Traceable to:	True Value Calibration Services Pvt. Ltd., Chennai				
Certificate No:	TVCSPL 23/09/991-01						
16. CALIBRATION RESULTS:							
Parameter: Volume							
Sl. No.	DUC Value	Standard value	Systematic Error	Permissible Systematic Error ±	Random Error	Permissible Random Error ±	Measurement Uncertainty±
	µl	µl	µl	µl	µl	µl	µl
1	11	10.90	0.10	0.2	0.07	0.1	0.16
2	30	29.84	0.16	0.5	0.09	0.2	0.16
3	50	49.80	0.20	0.5	0.08	0.2	0.16
17. REMARKS & CONCLUSION:							
<p>a. Reported Values of DUC are Average of 10 Measuring Series.</p> <p>b. The Measurement Uncertainty is estimated at a confidence level of 95.45 % with a coverage factor k=2.0.</p> <p>c. Calibration certificate issued for Pipette is used for scientific or industrial purpose only.</p> <p>d. Calibration of pipette is done as per ISO8655-6 & Permissible Systematic & Random error are given as per ISO 8655-2</p> <p>e. Coefficient of cubic thermal expansion (°C⁻¹ × 10⁻⁶) of Borosilicate Glass 5.0 is 15 × 10⁻⁶.</p> <p>f. The measurement data reported is "As found" - Without any Adjustment.</p> <p>g. This Certificate refer only to particular item submitted for Calibration.</p> <p>h. Volume at the reference temperature of 27°C, V27 from the apparent mass of water.</p> <p>i. The Std. used for calibration are calibrated and Traceable to National/International Standards.</p>							
Calibrated By				 <p>MANUNATH CHANDAKI DY TECHNICAL MANAGER Bengaluru</p>			
 NAGARAJ CALIBRATION ENGINEER							
End of certificate							
							FM-GL-67