







CALIBRATION CERTIFICATE

Page 1 of 1

1. Customer Name:	M/s. CHRONEVA HEALTH LAB PRIVATE LIMITED, The Orclid No-68, Ground Floor, 9th main, HMT Layout, RT Nagar, Bengaluru-560032.	6. Certificate No:	RCPL/23-24/ML/I5294-02				
		7. ULR No:	CC221523000014453F				
		8. Date of Issue:	02-11-2023				
2. SRF No:	I5294	9. Date of Calibration:	26-10-2023				
3. Date of SRF/Receipt:	18-10-2023	10. Next Cal Due:	25-10-2024				
4. Discipline:	Mechanical Calibration	11. Calibration SOP no:	RCPL/SOP/ML-21				
5. Calibrated At:	Mass & Volume Lab	12. Condition of DUC on Receipt:	Satisfactory				
13. ENVIRONMENTAL CONDITION : Temperature: 20.84 °C , Humidity: 58.33 %RH Ambient Pressure : : 911.5 hpa, Water Temperature: 20.2 °C							
14. DUC DETAILS:							
Nomenclature :	Micropipette	Serial No:	SW13151				
Make & Model :	Thermo Scientific & Finnppipette F3	Code/ID No :	CHVL_BLR_MP1_006				
Range :	5 - 50 µl	Resolution:	0.1 µl				
15. STANDARD INSTRUMENT USED:							
Nomenclature:	Electronic Semi Micro Balance	Serial No:	33604236				
Make / Model :	Sartorius / CPA225D	Cal Due Date:	10-09-2024				
Range :	0 to 220g	Traceable to:	True Value Calibration Services Pvt. Ltd., Chennai				
Certificate No:	TVCSPL 23/09/991-01						
16. CALIBRATION RESULTS:							
Parameter: Volume							
Sl. No.	DUC Value	Standard value	Systematic Error	Permissible Systematic Error ±	Random Error	Permissible Random Error ±	Measurement Uncertainty±
	µl	µl	µl	µl	µl	µl	µl
1	11	10.90	0.10	0.2	0.07	0.1	0.16
2	30	29.84	0.16	0.5	0.09	0.2	0.16
3	50	49.80	0.20	0.5	0.08	0.2	0.16
17. REMARKS & CONCLUSION:							
a. Reported Values of DUC are Average of 10 Measuring Series.							
b. The Measurement Uncertainty is estimated at a confidence level of 95.45 % with a coverage factor k=2.0.							
c. Calibration certificate issued for Pipette is used for scientific or industrial purpose only.							
d. Calibration of pipette is done as per ISO8655-6 & Permissible Systematic & Random error are given as per ISO 8655-2							
e. Coefficient of cubic thermal expansion (°C-1 × 10 ⁻⁶) of Borosilicate Glass 5.0 is 15 × 10 ⁻⁶ .							
f. The measurement data reported is "As found" - Without any Adjustment.							
g. This Certificate refer only to particular item submitted for Calibration.							
h. Volume at the reference temperature of 27°C, V27 from the apparent mass of water.							
i. The Std. used for calibration are calibrated and Traceable to National/International Standards.							
Calibrated By  NAGARAJ CALIBRATION ENGINEER				 Authorized Signatory MANJUNATH CHANDAKI DY TECHNICAL MANAGER Bengaluru			
End of certificate				FM-GL-67			

CALIBRATION CERTIFICATE




Page 1 of 1

1. Customer Name:	M/s. CHRONEVA HEALTH LAB PRIVATE LIMITED., The Orclid No-68, Ground Floor, 9th main, HMT Layout, RT Nagar, Bengaluru-560032.	6. Certificate No:	RCPL/23-24/ML/15294-03				
		7. ULR No:	CC221523000014454F				
		8. Date of Issue:	02-11-2023				
2. SRF No:	15294	9. Date of Calibration:	26-10-2023				
3. Date of SRF/Receipt:	15294	10. Next Cal Due:	25-10-2024				
4. Discipline:	Mechanical Calibration	11. Calibration SOP no:	RCPL/SOP/ML-21				
5. Calibrated At:	Mass & Volume Lab	12. Condition of DUC on Receipt:	Satisfactory				
13. ENVIRONMENTAL CONDITION : Temperature: 20.41 °C , Humidity: 54.88 %RH Ambient Pressure: : 910.2 hpa, Water Temperature: 20.5 °C							
14. DUC DETAILS:							
Nomenclature :	Micro Pipette	Serial. No:	RW01081				
Make & Model :	Thermo Scientific & Finnpipette F3	Code/ID No :	CHVL_BLR_MP2_007				
Range :	10 to 100 µl	Resolution:	0.2 µl				
15. STANDARD INSTRUMENT USED:							
Nomenclature:	Electronic Semi Micro Balance	Serial. No:	33604236				
Make / Model :	Sartorius / CPA225D	Cal Due Date:	10-09-2024				
Range :	0 to 220g	Traceable to:	True Value Calibration Services Pvt. Ltd., Chennai				
Certificate No:	TVCSP/09/991-01						
16. CALIBRATION RESULTS:							
Parameter: Volume							
Sl. No.	DUC Value	Standard value	Systematic Error	Permissible Systematic Error ±	Random Error	Permissible Random Error ±	Measurement Uncertainty±
	µl	µl	µl	µl	µl	µl	µl
1	11.0	11.00	0.00	0.10	0.06	0.08	0.16
2	50.0	49.81	0.19	0.50	0.03	0.20	0.16
3	100.0	100.64	-0.64	0.80	0.06	0.30	0.16
17. REMARKS & CONCLUSION:							
a. Reported Values of DUC are Average of 10 Measuring Series. b. The Measurement Uncertainty is estimated at a confidence level of 95.45 % with a coverage factor k=2.0. c. Calibration certificate issued for Pipette is used for scientific or industrial purpose only. d. Calibration of pipette is done as per ISO8655-6 & Permissible Systematic & Random error are given as per ISO 8655-2 e. Coefficient of cubic thermal expansion (°C ⁻¹ × 10 ⁻⁶) of Borosilicate Glass 5.0 is 15 × 10 ⁻⁶ . f. The measurement data reported is "As found" - Without any Adjustment. g. This Certificate refer only to particular item submitted for Calibration. h. Volume at the reference temperature of 27°C, V27 from the apparent mass of water. i. The Std. used for calibration are calibrated and Traceable to National/International Standards.							
Calibrated By  NAGARAJ CALIBRATION ENGINEER							
				 MANJUNATH CHANBAKI DY TECHNICAL MANAGER			
*** End Of Certificate ***							
FM-GL-67							

Note : 1) This Certificate refer only to particular item submitted for calibration. 2) The Calibration results reported are valid at the time of and under the stated conditions of the measurements. 3) The Calibration Certificate shall not be reproduced except in full without prior permission of the Laboratory. 4) The certificate of calibration shall not be utilized for any legal purpose in the court of law for which no responsibility will be associated to M/s. Reltec Calibration Pvt. Ltd. 5) Any error in the certificate should be brought to our lab within 45 days from the date of issue of certificate.

CALIBRATION CERTIFICATE

Page 1 of 1

1. Customer Name:	M/s. CHRONEVA HEALTH LAB PRIVATE LIMITED., The Orclid No-68, Ground Floor, 9th main, HMT Layout, RT Nagar, Bengaluru-560032.	6. Certificate No:	RCPL/23-24/ML/15294-01				
		7. ULR No:	CC221523000014452F				
		8. Date of Issue:	02-11-2023				
2. SRF No:	15294	9. Date of Calibration:	26-10-2023				
3. Date of SRF/Receipt:	26-10-2023	10. Next Cal Due:	25-10-2024				
4. Discipline:	Mechanical Calibration	11. Calibration SOP no:	RCPL/SOP/ML-21				
5. Calibrated At:	Mass & Volume Lab	12. Condition of DUC on Receipt:	Satisfactory				
13. ENVIRONMENTAL CONDITION: Temperature: 20.54 °C, Humidity: 55.80 %RH Ambient Pressure: : 910.4 hpa, Water Temperature: 20.3 °C							
14. DUC DETAILS:							
Nomenclature :	Micropipette	Serial No:	SW13594				
Make & Model :	Thermo Scientific & Finnpipette F3	Code/ID No :	CHVL_BLR_MP3_008				
Range :	100-1000 µl	Resolution:	1 µl				
15. STANDARD INSTRUMENT USED:							
Nomenclature:	Electronic Semi Micro Balance	Serial No:	33604236				
Make / Model :	Sartorius / CPA225D	Cal Due Date:	10-09-2024				
Range :	0 to 220g	Traceable to:	True Value Calibration Services Pvt. Ltd., Chennai				
Certificate No:	TVCSP1 23/09/991-01						
16. CALIBRATION RESULTS:							
Parameter: Volume							
Sl. No.	DUC Value	Standard value	Systematic Error	Permissible Systematic Error ±	Random Error	Permissible Random Error ±	Measurement Uncertainty±
	µl	µl	µl	µl	µl	µl	µl
1	100	100.66	-0.66	0.80	0.09	0.30	0.16
2	500	502.41	-2.41	4.00	0.26	1.50	0.46
3	1000	1005.74	-5.74	8.00	0.12	3.00	0.46
17. REMARKS & CONCLUSION:							
<p>a. Reported Values of DUC are Average of 10 Measuring Series.</p> <p>b. The Measurement Uncertainty is estimated at a confidence level of 95.45 % with a coverage factor k=2.0.</p> <p>c. Calibration certificate issued for Pipette is used for scientific or industrial purpose only.</p> <p>d. Calibration of pipette is done as per ISO8655-6 & Permissible Systematic & Random error are given as per ISO 8655-2</p> <p>e. Coefficient of cubic thermal expansion (°C-1 × 10⁻⁶) of Borosilicate Glass 5.0 is 15 * 10⁻⁶.</p> <p>f. The measurement data reported is "As found" - Without any Adjustment.</p> <p>g. This Certificate refer only to particular item submitted for Calibration.</p> <p>h. Volume at the reference temperature of 27°C, V27 from the apparent mass of water.</p> <p>i. The Std. used for calibration are calibrated and Traceable to National/International Standards.</p>							
Calibrated By				Authorized Signatory			
 NAGARAJ CALIBRATION ENGINEER							
				 MANJUNATH CHANDRANI DY TECHNICAL MANAGER			
				*** End of certificate***			
							FM-GL-67