

CALIBRATION CERTIFICATE

Page 1 of 2

1. Customer Name:	M/s. CHRONEVA HEALTH LAB PRIVATE LIMITED., The Orclid No-68, Ground Floor, 9th main, HMT Layout, RT Nagar, Bengaluru-560032.	6. Certificate No:	RCPL/23-24/ET/S2252-01
		7. ULR No.	CC221523000014648F
		8. Date of Issue:	03-11-2023
2. Discipline:	Electro Technical Calibration	9. Date of SRF:	26-10-2023
3. Date of Calibration:	26-10-2023	10. Next Cal Due:	25-10-2024
4. Calibrated At:	Oniste	11. Calibration SOP no:	RCPL/SOP/ET-05
5. SRF NO:	S2252	12. Condition of DUC on Receipt:	Functional
13. ENVIRONMENTAL CONDITION: Temperature: 20 to 40°C / Humidity: ≤70% R.H			
14. DUC DETAILS:			
Nomenclature :	Centrifuge (Timer)	Serial No:	FELC-6646
Make & Model :	REMI & R-8C	Code/ID No:	PDNBC_CF_001
Range :	0 to 60 min	Resolution:	1 min
Location :	Lab		
15. STANDARD INSTRUMENT USED:			
Nomenclature:	Time Totalizer	Serial. No:	105
Make & Model :	BELTRONICS & 501	Cal Due Date:	26-08-2024
Range :	As Per Manual	Traceable to:	Godrej & Boyce Mfg. Co. Ltd., Chennai.
Certificate No:	C-230821-14-1		
Calibrated By  PRASHANTH G CALIBRATION ENGINEER		Authorized Signatory  SRINIVAS TECHNICAL MANAGER Bengaluru	
		FM-GL-29	

CALIBRATION CERTIFICATE

Page 2 of 2

Certificate No	RCPL/23-24/ET/S2252-01	ULR NO.:	CC221523000014648F
----------------	------------------------	----------	--------------------

16. CALIBRATION RESULTS:

SI No	Parameter / Range	Standard Instrument Reading	Computed Value	DUC Reading	Computed Value	Error Obtained	Measurement Uncertainty \pm
		hr : min : sec : msec	sec	min	sec	sec	sec
1	Time / 0 to 60min	00:01:00:032	60.032	1.0	60.00	-0.032	0.054
2		00:05:00:036	300.036	5.0	300.00	-0.036	0.261
3		00:10:00:042	600.042	10.0	600.00	-0.042	0.522
4		00:20:00:043	1200.043	20.0	1200.00	-0.043	1.044
		hr : min : sec	sec	min	sec	sec	sec
5		00:30:00:051	1800.051	30.0	1800.00	-0.051	1.567

17. REMARKS & CONCLUSION:

- Reported Values of Standard & DUC are Average of 5 Trials.
- The Measurement Uncertainty Values are estimated at a confidence level of 95.45 % with a coverage factor $k = 2.00$.
- Calibration points & Due date is given as per customer request.
- Measurement results are traceable to national standard.

Calibrated By


PRASHANTH G

CALIBRATION ENGINEER

Authorized Signatory


B. PRINIVAS
TECHNICAL MANAGER

End of Certificate

FM-GL-29