

CALIBRATION CERTIFICATE

DATE: 17-11-2023

NAME OF CUSTOMER:

KG CLINICAL LABORATORY
4/526, PUTHENPURACKAL BUILDING
THEATRE JUNCTION ANAKKARE
IDUKK, KERALA

MISPA NEO CHEMISTRY SEMI AUTO ANALYZER PHOTOMETER BEARING.

S/NO: WR-47015667CALIBRATIONS ON: 17-11-2023

CALIBRATION VALUES:

LIGHT INTENSITY TEST

FILTERS	GAIN	BACKGROUND	DARK CURRENT
340		57283	94
405	45	57248	80
450	51	57360	75
510	55	58256	81
546	57	58214	79
578	59	57158	80
630	58	57418	86
670	61	58479	83

PERISTALTIC PUMP ADJUSTMENT TEST

<u>VALUES</u>	<u>CURRENT</u>	<u>NEW</u>
<u>SAMPLE VOLUME</u>	<u>44225</u>	<u>44225</u>
<u>POSITIONING</u>	<u>1502</u>	<u>1802</u>
<u>PUMB DELAY</u>	<u>2</u>	<u>2</u>

NEXT CALIBRATION DUE ON: 16.11.2024

THIS INSTRUMENT IS CALIBRATED BY USING THE PERFORMANCE OF THE LIQUID CALIBRATOR ONLY

TESTED BY: G DINESH

APPROVED BY: Q.A



NOTE: 1) Calibration Points & due date is given as per customer request 2) This certificate refer only to particular item submitted for certificate 3) The calibration results reported are valid at the time of and under the stated conditions of the measurements 4) Any error in the certificate should be brought to our lab within 45 days from the date of issue of certificate.