

MK BEST CALIBRATION SERVICES

NABL ACCREDITED CALIBRATION LABORATORY AS PER ISO/IEC17025 : 2017

No. 27, F-2, 1st Floor, 2nd Street, Varalakshmi Nagar, Maduravoyal,
(Opp. MGR Engineering College), Chennai - 600 095.

Ph.: 044 - 23780211, Cell : 93802 66480 / 86958 18108 / 90032 77250

E-mail: mkbestcalibration@gmail.com, www.mkbestcalibrationservices.com



CC-3340



CERTIFICATE OF CALIBRATION

FF/7.8/01

1 of 1

ULR No	CC334023000008207F	Date of Calibration	31.10.2023	Date of Receipt	31.10.2023
Certificate No	MKBL/23/10/0941-002	Due Date of Calibration	30.10.2024	Date of Issue	02.11.2023

CUSTOMER INFORMATION

M/S. VJA LABORATORY,
THIRUKOILUR MAIN ROAD,
NEAR VADAKURUMBUR ROAD,
ERAIYUR, VILLUPURAM - 605 602.

DETAILS OF UNIT UNDER CALIBRATION

Description	MICROPIPETTE 2
Make / Model	CORAL / MK431291
Range/Resolution	2 to 20 µl / 1 µl
Serial No	VV100
Identification No	JDC/CB/GEN/03
Location	-----
Calibrated at	LAB

STANDARD INSTRUMENTS DETAILS (The Standards Used are Traceable to National /International Standards)

S.No	Description	Id.No/Sl. No	Certificate No	Validity
01	ELECTRONIC SEMI MICRO BALANCE	MK/CAL-96/477904	TVCSPL 23/07/638-01	20.07.2024
02	DIGITAL BAROMETER	MK/CAL-143	CC287923000005505F	04.10.2024
03	DIGITAL PEN TYPE THERMOMETER	MK/CAL-160	CRMTL/01/423101578-A4	03.07.2024

ENVIRONMENTAL & DUC CONDITIONS

REFERENCE STANDARD

Temperature	23.6 °C	Reference Std	ISO 8655-6:2002
Humidity	54 % RH	Procedure No	MKBCS - MBV - 03
Condition of UUC Receipt	Good		

CALIBRATION RESULTS

L.VOLUME CALIBRATION

S.No	UUC Reading (Mean) µl	STD Reading (Mean) µl	Deviation µl	Expanded Uncertainty (±) µl
1	2	2.00	0.00	1.18
2	5	5.01	-0.01	1.18
3	10	10.01	-0.01	1.18
4	15	15.03	-0.03	1.18
5	20	20.03	-0.03	1.18

Remarks :

- The Expanded Uncertainty Associated with the Results is Calculated at a Confidence Level of Approximately 95% with a Coverage factor of K=2.
- The Calibration Certificate Shall not be Reproduced Except In Full, Without Written Approval Of The Laboratory.
- The Recalibration Interval Should be Determined on the User Requirement.
- The Results Stated In This Certificate Relate Only to the Item Calibrated.
- The User Should Determine The Suitability Of The Instrument For Its Intended Use.
- Resulted Volume Convert at 27°C of Water Temperature.
- Expanded Uncertainty is also Included Correction Factors.

Calibrated by

S. Chandrabose
S.Chandrabose
(Calibration Engineer)



Authorised By

L. Magesh
L.Magesh
(MD/QM)