

MK BEST CALIBRATION SERVICES



No. 27, F-2, 1st Floor, 2nd Street, Varalakshmi Nagar, Maduravoyal,
(Opp. MGR Engineering College), Chennai - 600 095.

Ph.: 044 - 23780211, Cell : 93802 66480 / 86958 18108 / 90032 77250

E-mail: mkbestcalibration@gmail.com, www.mkbestcalibrationservices.com



CERTIFICATE OF CALIBRATION



FF/7.8/01

Certificate No	: MKBN/23/10/0515-001	Date of Issue	: 02.11.2023
Date of Calibration	: 31.10.2023	Due Date of Calibration	: 30.10.2024

CUSTOMER INFORMATION		UNIT UNDER CALIBRATION	
M/S. VJA LABORATORY., THIRUKOILUR MAIN ROAD, NEAR VADAKURUMBUR ROAD, ERAIYUR, VILLUPURAM - 605 602.		Description	CENTRIFUGE
		Range	3600 rpm
		Resolution	100 rpm
		Make	REMI
		Model	C 854/4
		Identification No	VJA/CB/GEN/01
		Serial No	BHLC-2080
		Location	LABORATORY

S.No	Description	Id.No/Sl No	Traceability	Certificate No	Validity
01	Digital Tachometer	MK/CAL-222	National Standards	CRMTL/01/422103978-A4	02.12.2023

ENVIRONMENTAL CONDITIONS		REFERENCE STANDARD	
Temperature	23.6°C	Uncertainty ± °C	1.57
Humidity	54 % RH		

CALIBRATION RESULTS			
S.NO	Standard Reading rpm	UUC Readings rpm	Error rpm
1	505	500	-5
2	1014	1000	-14
3	2039	2000	-39
4	3048	3000	-48
5	3665	3600	-65

Remarks :

1. The above UUC was Calibrated at Site.
2. The Calibration Certificate Shall not be Reproduced Except In Full, Without Written Approval Of The Laboratory.
3. The Recalibration Interval is Determined based on the User requirement.
4. The Results Stated In This Certificate Relate Only to the Item Calibrated.
5. The above Instrument used for Medical Purpose.

Calibrated by

S. Chandrabose
S.Chandrabose
(Calibration Engineer)

X-X-X-X- End Of Certificate -X-X-X-X



Authorised by

L. Magesh
L.Magesh
(MD/QM)