

VIEW LAB MONTHLY SUMMARY

Lab Name	ISAAC CLINICAL LABORATORY		Details About Robust Analysis							
Lab No	11844		Detail About Monthly Summary							
Month	May		Detail about SDI							
Year	2023									
Constituent Group	Chemistry II									
Click on the analyte to view Graphical Data			All Analyser Result		Print					
			Date of Result Entered : 26/05/2023							
			Date of Report Published : 05/06/2023							
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE II	GOD-POD II	Biosystems	310	186.96	8.32	15.56	189.5 mg/dL	0.16	1.77
2	UREA II	Urease UV / GLDH II	Biosystems	224	21.96	16.54	3.63	24 mg/dL	0.56	0.49
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Biosystems	236	6.66	12.14	0.81	7.15 mg/dL	0.61	0.11
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Biosystems	230	2.39	15.66	0.37	2.28 mg/dL	-0.29	0.05
5	T-PROTEIN II	Biuret - Colorimetric II	Biosystems	278	5.78	9.57	0.55	6.18 g/dL	0.72	0.07
6	ALBUMIN II	BCG - Colorimetric II	Biosystems	268	3.38	9.52	0.32	3.26 g/dL	-0.37	0.04
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Biosystems	260	6.94	11.17	0.78	7.23 mg/dL	0.37	0.10
8	CHOLESTROL II	CHOD-PAP II	Biosystems	289	121.21	9.83	11.91	118.5 mg/dL	-0.23	1.40
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Biosystems	268	132.35	11.45	15.15	136.5 mg/dL	0.27	1.85
SDI Range			Interpretation							
Within -1.00 to +1.00			Excellent.							
Within ±1.01 to ±2.00			Good.							
Within ±2.01 to ±2.99			Accept with caution. Warning Signal.							
Beyond ±3.0			Unacceptable performance. Action Signal.							

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

