

Date: 18/08/2023

CALIBRATION CERTIFICATE

This is to certify that **BTS350** Photometer bearing **S. No: 801759758** installed at **Issac Clinical Laboratory –Tirukovilur** has been calibrated on **18/08/2023**

Calibration Values:

LIGHT INTENSITY TEST

LED	TINT	MAIN	REFERENCE
340	500	970341	283836
405	129	946545	702960
505	420	976201	724000
535	157	986667	545436
560	114	988403	723330
600	500	959902	663294
635	120	944170	552560
670	13	951546	762058

PERISTALTIC PUMP ADJUSTMENT TEST

VALUE	CURRENT	NEW
SAMPLE VOLUME	43574	43574
POSITIONING	2200	2200
PUMP DELAY	2	2

Next calibration is due on **17/08/2024**

For **BIOSYSTEMS DIAGNOSTICS PVT. LTD.,**

KRUNAL
GORDHANBHAI
MADHU

Digitally signed by KRUNAL
GORDHANBHAI MADHU
DN: cn=KRUNAL GORDHANBHAI
MADHU, c=IN, o=PERSONAL,
email=krunalmadhu@biosystems.in
Date: 2023.10.10 17:36:47 +05'30'

AUTHORISED SIGNATORY