

# MK BEST CALIBRATION SERVICES

NABL ACCREDITED CALIBRATION LABORATORY AS PER ISO/IEC17025 : 2017



No. 27, F-2, 1st Floor, 2nd Street, Varalakshmi Nagar, Maduravoyal,  
(Opp. MGR Engineering College), Chennai - 600 095.

Ph.: 044 - 23780211, Cell : 93802 66480 / 86958 18108 / 90032 77250

E-mail: mkbestcalibration@gmail.com, www.mkbestcalibrationservices.com



CC-3340

## CERTIFICATE OF CALIBRATION

FF/7.8/01		Page No		1 of 1	
ULR No	CC334023000007799F	Date of Calibration	21.09.2023	Date of Receipt	21.09.2023
Certificate No	MKBL/23/09/0866-002	Due Date of Calibration	20.09.2024	Date of Issue	23.09.2023

CUSTOMER INFORMATION	DETAILS OF UNIT UNDER CALIBRATION
M/S. ISAAC CLINICAL LABORATORY., NO.8 A, ESAT STREET, TIRUKOVILUR, KALLAKURICHI DISTRICT - 605757.	Description <b>MICROPIPETTE 2</b>
	Make / Model <b>FINN PIPETTE / 4640060</b>
	Range/Resolution <b>100 to 1000 µl / 100 µl</b>
	Serial No <b>TW11905</b>
	Identification No <b>ICL/CB/GEN/03</b>
	Location <b>-----</b>
	Calibrated at <b>LAB</b>

STANDARD INSTRUMENTS DETAILS (The Standards Used are Traceable to National /International Standards)				
S.No	Description	Id.No/Sl. No	Certificate No	Validity
01	ELECTRONIC SEMI MICRO BALANCE	MK/CAL-96/477904	TVCSP/23/07/638-01	20.07.2024
02	DIGITAL BAROMETER	MK/CAL-143	CC287922000005116F	12.10.2023
03	DIGITAL PEN TYPE THERMOMETER	MK/CAL-160	CRMTL/01/42310157B-A4	03.07.2024

ENVIRONMENTAL & DUC CONDITIONS		REFERENCE STANDARD	
Temperature	23.5 °C	Reference Std	ISO 8655-6:2002
Humidity	53 % RH	Procedure No	MKBCS - MBV - 03
Condition of UUC Receipt	Good		

### CALIBRATION RESULTS

#### I.VOLUME CALIBRATION

S.No	UUC Reading (Mean) µl	STD Reading (Mean) µl	Deviation µl	Expanded Uncertainty (±) µl
1	200	200.00	0.00	1.18
2	400	400.01	-0.01	1.18
3	600	600.01	-0.01	1.18
4	800	800.02	-0.02	1.18
5	1000	1000.02	-0.02	1.18

- Remarks :**
- The Expanded Uncertainty Associated with the Results is Calculated at a Confidence Level of Approximately 95% with a Coverage factor of K=2.
  - The Calibration Certificate Shall not be Reproduced Except In Full, Without Written Approval Of The Laboratory.
  - The Recalibration Interval Should be Determined on the User Requirement.
  - The Results Stated In This Certificate Relate Only to the Item Calibrated.
  - The User Should Determine The Suitability Of The Instrument For Its Intended Use.
  - Resulted Volume Convert at 27°C of Water Temperature.
  - Expanded Uncertainty is also Included Correction Factors.

Calibrated by  
  
S. Chandrabose  
(Calibration Engineer)



Authorised By  
  
L. Magesh  
(MD/QM)