

# MK BEST CALIBRATION SERVICES



No. 27, F-2, 1st Floor, 2nd Street, Varalakshmi Nagar, Maduravoyal,  
(Opp. MGR Engineering College), Chennai - 600 095.

Ph.: 044 - 23780211, Cell : 93802 66480 / 86958 18108 / 90032 77250

E-mail: mkbestcalibration@gmail.com, www.mkbestcalibrationservices.com



## CERTIFICATE OF CALIBRATION



FF/7.8/01

<b>Certificate No</b> :	MKBN/23/09/0401-001	<b>Date of Issue</b> :	23.09.2023
<b>Date of Calibration</b> :	21.09.2023	<b>Due Date of Calibration</b> :	20.09.2024

CUSTOMER INFORMATION		UNIT UNDER CALIBRATION	
M/S. ISAAC CLINICAL LABORATORY., NO.8 A, ESAT STREET, TIRUKOVILUR, KALLAKURICHI DISTRICT - 605757.		<b>Description</b>	CENTRIFUGE
		<b>Range</b>	3600 rpm
		<b>Resolution</b>	100 rpm
		<b>Make</b>	REMI
		<b>Model</b>	C-852
		<b>Identification No</b>	ICL/CB/GEN/01
		<b>Serial No</b>	ZHKN-43546
		<b>Location</b>	LAB

S.No	Description	Id.No/SL No	Traceability	Certificate No	Validity
01	Digital Tachometer	MK/CAL-222	National Standards	CRMTL/01/422103978-A4	02.12.2023

ENVIRONMENTAL CONDITIONS		REFERENCE STANDARD	
Temperature	24.2°C	Uncertainty ± °C	1.57
Humidity	53 % RH		

### CALIBRATION RESULTS

S.NO	Standard Reading rpm	UUC Readings rpm	Error rpm
1	509	500	-9
2	1015	1000	-15
3	2021	2000	-21
4	3032	3000	-32
5	3639	3600	-39

#### Remarks :

- The above UUC was Calibrated at Site.
- The Calibration Certificate Shall not be Reproduced Except In Full, Without Written Approval Of The Laboratory.
- The Recalibration Interval is Determined based on the User requirement.
- The Results Stated In This Certificate Relate Only to the Item Calibrated.
- The above Instrument used for Medical Purpose.

Calibrated by

*S. Chandrabose*

S.Chandrabose  
(Calibration Engineer)

x-x-x-x- End Of Certificate -x-x-x-x



Authorised by

*L. Magesh*

L.Magesh  
(MD/QM)