

Calibration Certificate

CC No: 1010/ 23

Date: 13/06/2023

Customer Name:

VJA Laboratory

Address:

Thirukovilur Main Road, Eraiyur-607201 .

Name of the Instrument & Model	SEMI AUTO BIOCHEMISTRY ANALYZER & SA2000
--------------------------------------	---

Calibration Table

Power Supply (Main)	Actual Reading in volts	Reading Shown in Volts	Remarks
Phase - Neutral	230	228	All Readings are within range. Instrument is working fine
Phase - Earth	230	227	
Earth - Neutral	0	1	
Instrument power supply (DC)	12	12.1	
Lamp Voltage	6	6	
Photo Meter Reading			
Filters In NM	Actual OD in ABS	OD Shown in ABS	
340	0.5	0.502	
405	0.5	0.501	
492	0.5	0.501	
510	0.5	0.502	
546	0.5	0.498	
578	0.5	0.497	
630	0.5	0.501	

Next calibration due on:

12/06/2024

Instrument Status:

Calibration done. Working Fine

Calibrated by *C. Karuppusamy*
Krystal Medicare
 Sign: *(C. Karuppusamy)*
DIAGNOSTIC & SURGICALS DEALERS

New No.82, T.K. Mudali Street, Choolai, Chennai - 600112, South India
 Ph: 044-42102622 Cell: 9840140283 - E mail : krystal2021medicare@gmail.com

