

VIEW LAB MONTHLY SUMMARY

Lab Name	VJA LABORATORY	Details About Robust Analysis								
Lab No	11859	Detail About Monthly Summary								
Month	October	Detail about SDI								
Year	2023									
Constituent Group	Chemistry II									
		Click on the analyte to view Graphical Data			All Analyser Result		Print			
		Date of Result Entered : 21/10/2023								
		Date of Report Published : 04/11/2023								
Sl.No	Analyte	Method / Principle Name	Analyser Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	1673	129.03	9.72	12.54	110 mg/dL	-1.52	0.61
2	UREA II	Urease UV / GLDH II	Any Analyser (Automation / Semi Automation)	1043	86.31	11.17	9.64	85.2 mg/dL	-0.12	0.60
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Any Analyser (Automation / Semi Automation)	1204	7.35	12.30	0.90	6.9 mg/dL	-0.50	0.05
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Any Analyser (Automation / Semi Automation)	1325	8.35	11.84	0.99	7.5 mg/dL	-0.86	0.05
5	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	1282	5.47	10.35	0.57	4.8 g/dL	-1.18	0.03
6	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	1253	3.29	9.69	0.32	3.1 g/dL	-0.60	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	1268	5.07	16.36	0.83	4.9 mg/dL	-0.21	0.05
8	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	1485	109.78	11.30	12.40	107 mg/dL	-0.22	0.64
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	1243	104.67	16.11	16.86	73 mg/dL	-1.88	0.96
SDI Range		Interpretation								
Within -1.00 to +1.00		Excellent.								
Within ±1.01 to ±2.00		Good.								
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

