

## VIEW LAB MONTHLY SUMMARY

Lab Name	SRI THANVANDARI CLINICAL LAB		<a href="#">Details About Robust Analysis</a>							
Lab No	16792		<a href="#">Detail About Monthly Summary</a>							
Month	October		<a href="#">Detail about SDI</a>							
Year	2023									
Constituent Group	Chemistry II									
<a href="#">Click on the analyte to view Graphical Data</a>			<a href="#">All Analyser Result</a>		<a href="#">Print</a>					
			Date of Result Entered : 20/10/2023							
			Date of Report Published : 04/11/2023							
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	<a href="#">GLUCOSE II</a>	GOD-POD II	Any Analyser (Automation / Semi Automation )	1673	129.03	9.72	12.54	121.01 mg/dL	-0.64	0.61
2	<a href="#">UREA II</a>	Urease UV / GLDH II	Any Analyser (Automation / Semi Automation )	1043	86.31	11.17	9.64	113.12 mg/dL	2.78	0.60
3	<a href="#">CREATININE II</a>	Jaffes Kinetic-Alkaline Picrate II	Any Analyser (Automation / Semi Automation )	1204	7.35	12.30	0.90	7.18 mg/dL	-0.19	0.05
4	<a href="#">T.BILIRUBIN II</a>	Diazonium Salt ( Colorimetric ) / Jendrassik II	Any Analyser (Automation / Semi Automation )	1325	8.35	11.84	0.99	7.96 mg/dL	-0.39	0.05
5	<a href="#">T-PROTEIN II</a>	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation )	1282	5.47	10.35	0.57	5.4 g/dL	-0.12	0.03
6	<a href="#">ALBUMIN II</a>	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation )	1253	3.29	9.69	0.32	2.89 g/dL	-1.25	0.02
7	<a href="#">URIC ACID II</a>	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation )	1268	5.07	16.36	0.83	5.53 mg/dL	0.55	0.05
8	<a href="#">CHOLESTROL II</a>	CHOD-PAP II	Any Analyser (Automation / Semi Automation )	1485	109.78	11.30	12.40	126.31 mg/dL	1.33	0.64
9	<a href="#">TRIGLYCERIDE II</a>	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation )	1243	104.67	16.11	16.86	122.73 mg/dL	1.07	0.96
<b>SDI Range</b>			<b>Interpretation</b>							
Within -1.00 to +1.00			Excellent.							
Within ±1.01 to ±2.00			Good.							
Within ±2.01 to ±2.99			Accept with caution. Warning Signal.							
Beyond ±3.0			Unacceptable performance. Action Signal.							

## Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>

