



Date: 26-10-2023  
Effective Date: 26-10-2023

## Certificate of Inspection

1. Model: Automated Hematology Analyzer H560
2. Serial No.: K11042137031
3. Calibration Date: 26-10-2023
4. Material used: H Cal (Lot No. PLUS1023, Expiry date: 10-Nov-2023)

By comparing your data to the results of the standard counters in Erba Lachema, the calibration for CBC 5 parameters using the measurement standard material (H Cal) was completed. The calibration result of 5 runs is summarized in the following table. Please refer to the attached sheets for the details.



Technical Service Department  
Transasia Bio-Medicals Ltd



Date: 26-10-2023  
Effective Date: 26-10-2023

## CERTIFICATE OF CALIBRATION

**Customer Name:** CHRONEVA HEALTH LAB PRIVATE LIMITED

**Model :** Automated Hematology Analyzer H560


**Serial No. :** K11042137031

**Calibration Done Date:** 26-10-2023

**Next Calibration Due Date On or Before:** 25-10-2024

**Lab In-charge: .** DINAKARAN

*This is to certify that the above-mentioned product has been verified of calibration for CBC 5 parameters (WBC, RBC, HGB, MCV and PLT) according to the standard procedures provided by Erba Lachema s.r.o, Karasek.*



Calibration at site performed by  
Engineer Name ATHISH M  
Designation SERVICE ENGINEER  
Transasia Bio-Medicals Ltd  
Location

Encl:

1. Certificate of Inspection
2. Assay Sheet of Hematology Calibrator (H Cal)
3. Printouts
4. Traceability Document



Date: 26-10-2023  
Effective Date: 26-10-2023

## Certificate of Inspection

1. Model: Automated Hematology Analyzer H560
2. Serial No.: K11042137031
3. Calibration Date: 26-10-2023
4. Material used: H Cal (Lot No. PLUS1023, Expiry date: 10-Nov-2023)

By comparing your data to the results of the standard counters in Erba Lachema, the calibration for CBC 5 parameters using the measurement standard material (H Cal) was completed. The calibration result of 5 runs is summarized in the following table. Please refer to the attached sheets for the details.



Technical Service Department  
Transasia Bio-Medicals Ltd



UNMATCHED SERVICE  
SINCE 1979...

## 5. BACKGROUND CHECK

PARAMETER	RESULT	Range
WBC	0.0	$0.3 \times 10^3$ /UI or Less
RBC	0.00	$0.02 \times 10^9$ /uL or Less
HGB	0.0	0.1 g/dL or Less
PLT	0	$10 \times 10^3$ /uL or Less

Technical Service Department  
Transasia Bio-Medicals Ltd



## 6. PRECISION STUDY PERFORMED ON THE ANALYSER USING A BLOOD SAMPLE (ORIGINALS ATTACHED)

SMP NO	WBC	RBC	HGB	MCV	PLT
1	7.12	4.87	11.3	72.9	262
2	6.94	4.88	11.3	73.1	270
3	7.28	4.95	11.3	73.1	276
4	7.14	4.92	11.3	73.0	262
5	7.14	4.88	11.2	73.0	271
6	7.08	4.82	11.3	73.0	262
7	7.08	4.92	11.2	73.0	270
8	7.31	4.87	11.2	72.9	270
9	7.23	4.93	11.3	73	256
10	6.97	4.93	11.3	72.8	273
Mean	7.13	4.90	11.27	72.98	267.20
SD	0.12	0.04	0.05	0.09	6.29
<b>CV%</b>	<b>1.70</b>	<b>0.81</b>	<b>0.43</b>	<b>0.13</b>	<b>2.35</b>
Acceptable CV%	Within 3.5%	Within 2.0%	Within 1.5%	Within 2.0%	Within 6.0%
<b>Result</b>	<b>PASS</b>	<b>PASS</b>	<b>PASS</b>	<b>PASS</b>	<b>PASS</b>

  
 Technical Service Department  
 Transasia Bio Medicals Ltd