

CALIBRATION CERTIFICATE

Page 1 of 1

1. Customer Name:	M/s. CHRONEVA HEALTH LAB PRIVATE LIMITED, The Orclid No-68, Ground Floor, 9th main, HMT Layout, RT Nagar, Bengaluru-560032	6. Certificate No:	RCPL/23-24/ML/S2252-02
		7. ULR No:	CC221523000014511F
		8. Date of Issue:	03-11-2023
2. SRF NO:	S2252	9. Date of Calibration:	26-10-2023
3. Date of SRF/Receipt:	26-10-2023	10. Next Cal Due:	25-10-2024
4. Discipline:	Mechanical Calibration	11. Calibration SOP no:	RCPL/SOP/ML-05
5. Calibrated At:	Onsite	12. Condition of DUC on Receipt:	Functional

13. ENVIRONMENTAL CONDITION : Temperature: 23.4°C / Humidity: 54% R.H

14. DUC DETAILS:

Nomenclature :	Centrifuge (RPM)	Serial. No:	FFLC-6646
Make / Model :	Remi & R-8C	Code/ID No :	CHVL_BLR_C_003
Range :	Upto 5250 RPM	Resolution:	10 RPM
Location:	LAB		

15. STANDARD INSTRUMENT USED:

Nomenclature:	Digital Tachometer	Serial. No:	164-0109
Make & Model :	Line Seiki & TM-5000	Cal Due Date:	17-09-2024
Range :	6.0 to 99999.9 RPM	Traceable to:	Sansel Instruments and Control, Chennai
Certificate No:	SANS/FT/01(7.8/01)		

16. CALIBRATION RESULTS:


Parameter: SPEED

Sl. No.	Standard Reading (RPM)	DUC Set Point (RPM)	Error Observed (RPM)	Uncertainty of Measurement ± (% of reading)
1	500.3	500	-0.3	1.16
2	1000.3	1000	-0.3	0.58
3	2000.7	2000	-0.7	0.29
4	4000.7	4000	-0.7	0.15
5	5200.7	5250	49.3	0.11

17. REMARKS & CONCLUSION:

- Reported Values of DUC & Standard are Average of 5 Measuring Series.
- The Measurement Uncertainty is estimated at a confidence level of 95.45 % with a coverage factor $k=2.0$.
- Calibration certificate issued for Centrifuge is used scientific or industrial purpose only.
- The Std. used for calibration were calibrated and Traceable to National/International Standards.

Calibrated By



SHRIKANTH BIRADAR
CALIBRATION ENGINEER



****End Of Certificate****

Authorized Signatory



MANJUNATH CHANDAKI
DY TECHNICAL MANAGER
Bengaluru

FM-GL-85

CALIBRATION CERTIFICATE

Page 1 of 2

1. Customer Name:	M/s. CHRONEVA HEALTH LAB PRIVATE LIMITED., The Orclid No-68, Ground Floor, 9th main, HMT Layout, RT Nagar, Bengaluru-560032.	6. Certificate No:	RCPL/23-24/ET/S2252-01
		7. ULR No.	CC221523000014648F
		8. Date of Issue:	03-11-2023
2. Discipline:	Electro Technical Calibration	9. Date of SRF:	26-10-2023
3. Date of Calibration:	26-10-2023	10. Next Cal Due:	25-10-2024
4. Calibrated At:	Oniste	11. Calibration SOP no:	RCPL/SOP/ET-05
5. SRF NO:	S2252	12. Condition of DUC on Receipt:	Functional

13. ENVIRONMENTAL CONDITION: Temperature: 20 to 40°C / Humidity: ≤70% R.H

14. DUC DETAILS:

Nomenclature :	Centrifuge (Timer)	Serial No:	FELC-6646
Make & Model :	REMI & R-8C	Code/ID No:	PDNBC_CF_001
Range :	0 to 60 min	Resolution:	1 min
Location :	Lab		

15. STANDARD INSTRUMENT USED:

Nomenclature:	Time Totalizer	Serial. No:	105
Make & Model :	BELTRONICS & 501	Cal Due Date:	26-08-2024
Range :	As Per Manual	Traceable to:	Godrej & Boyce Mfg. Co. Ltd., Chennai.
Certificate No:	C-230821-14-1		


Calibrated By


PRASHANTH G

CALIBRATION ENGINEER



Authorized Signatory


SRINIVAS
TECHNICAL MANAGER
Bangalore

FM-GL-29

