

## CALIBRATION CERTIFICATE

Page 1 of 1

1. Customer Name:	M/s. CHRONEVA HEALTH LAB PRIVATE LIMITED., The Orclid No-68, Ground Floor, 9th main, HMT Layout, RT Nagar, Bengaluru-560032.	6. Certificate No:	RCPL/23-24/TH/S2252-01	
		7.ULR:	CC221523000014575F	
		8.Date of ISSUE:	02-11-2023	
2.SRF NO:	S2252	9. Date of Calibration:	26-10-2023	
3.Date of SRF/Receipt:	26-10-2023	10. Next Cal Due:	25-10-2024	
4.Discipline:	Thermal Calibration	11. Calibration SOP no:	RCPL/SOP/TH-09	
5.Calibrated At:	Onsite	12. Condition of DUC on Receipt:	Functional	
13.ENVIRONMENTAL CONDITION: Temperature: 20 to 40°C / Humidity: ≤70%R.H				
14. DUC DETAILS:				
Nomenclature :	Digital Temperature Controller with Sensor of Refrigerator	Serial No:	03944ZAQ802655N	
Make & Model :	Samsung & RA18/2008	ID No:	CHVL_BLR_RF1_009	
Range :	2 to 8°C	Resolution:	1°C	
Location :	Lab			
15. STANDARD INSTRUMENT USED:				
Nomenclature:	RTD WITH INDICATOR	Serial No:	Sen: 1591, Ind: 14F237012	
Make & Model :	TEMPSENS & TEMPMET-08	Cal Due Date:	18-09-2024	
Range :	-80 to 400°C	Traceable to:	TEMPSENS, Udaipur	
Certificate No:	TL/023/1088.1.2			
16.CALIBRATION RESULTS:				
Sl.No	Standard Meter Reading (°C)	DUC Set (°C)	Error Obtained (°C)	Expanded Measurement Uncertainty ± (°C)
1	5.72	5	-0.72	0.59
17.REMARKS & CONCLUSION:				
<p>a. Reported Values of STD &amp; DUC are Average of 5 Trials.</p> <p>b. The Measurement Uncertainty is estimated at a confidence level of 95.45 % with a coverage factor k=2.00.</p> <p>c. Calibration Points &amp; Due date is given as per Customer Request.</p> <p>d. Temperature Scale: International Temperature Scale-1990(ITS-90)</p> <p>e. The Std.used for calibration were calibrated and Traceable to National/InternationalStandards.</p> <p>f. Use of NABL Symbol are adopted as per guidelines of NABL-133.</p>				
Calibrated By			Authorized Signatory	
 PRASHANTH G CALIBRATION ENGINEER			 DY.TECHNICAL MANAGER	
		***End of Certificate***		