




CALIBRATION CERTIFICATE

Page 1 of 1

1. Customer Name:	M/s. CHRONEVA HEALTH LAB PRIVATE LIMITED., The Orclid No-68, Ground Floor, 9th main, HMT Layout, RT Nagar, Bengaluru-560032.	6. Certificate No:	RCPL/23-24/TH/15294-01		
		7. ULR No:	CC221523000014323F		
		8. Date of ISSUE:	30-10-2023		
2. SRF No:	15294	9. Date of Calibration:	26-10-2023		
3. Date of SRI/Receipt:	26-10-2023	10. Next Cal Due:	25-10-2024		
4. Discipline:	Thermal Calibration	11. Calibration SOP No:	RCPL/SOP/TH-13		
5. Calibrated At:	Thermal Lab	12. Condition of DUC on Receipt:	Functional		
13. ENVIRONMENTAL CONDITION: Temperature: 25±4°C / Humidity: 30 to 75% R.H					
14. DUC DETAILS:					
Nomenclature :	Digital Thermo Hygrometer	Serial No:	11451118		
Make & Model :	HTC & HTC-1	ID No:	CHVL_BLR_THI_014		
Range :	-50 to 70°C & 10 to 99%R.H.	Resolution:	0.1° C & 1% R.H.		
Accuracy :	±1°C & ±5% RH				
15. STANDARD INSTRUMENT USED:					
1. Nomenclature:	Digital Temp & Humidity Meter	Serial No:	IND:5210392, SEN:20496567		
Make & Model :	Rotronic & HP-32	Cal Due Date:	27-11-2023		
Range :	-10 to 60°C/ 0 to 100%RH	Traceable to:	HI TECH, Chennai		
Certificate No:	HT/CC/221128-20/001				
16. CALIBRATION RESULTS:					
Parameter: Temperature & Humidity					
Sl No.	Set Value	Standard Meter Reading	DUC Reading	Error Obtained	Measurement Uncertainty ±
Temperature(°C) @50%RH					
	°C	°C	°C	°C	°C
1	15	15.12	15.0	-0.12	0.24
2	30	30.07	29.8	-0.27	0.24
3	50	50.14	49.8	-0.34	0.24
Humidity (%RH) @25°C					
	%RH	%RH	%RH	%RH	%RH
4	30	30.11	29	-1.11	1.17
5	60	59.19	58	-1.19	1.17
6	95	92.25	91	-1.25	1.17
17. REMARKS & CONCLUSION:					
a. Reported Values of STD & DUC are Average of 5 Trials.					
b. The Measurement Uncertainty is estimated at a confidence level of 95.45 % with a coverage factor k=2.00.					
c. Calibration Points & Due date is given as per Customer Request.					
d. The Std. used for calibration were calibrated and Traceable to National/International Standards.					
Calibrated By				 BEERAPPA DY. TECHNICAL MANAGER	
 KAVYA RAJ.K. CALIBRATION ENGINEER					
*** End of Certificate ***					
FM-GL-29					