

VIEW LAB MONTHLY SUMMARY

Lab Name	JAYAM CLINICAL LABORATORY	Details About Robust Analysis									
Lab No	11843	Detail About Monthly Summary									
Month	May	Detail about SDI									
Year	2023										
Constituent Group	Chemistry II										
		Click on the analyte to view Graphical Data				All Analyser Result		Print			
		Date of Result Entered : 19/05/2023									
		Date of Report Published : 05/06/2023									
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	Participants SD	Your Value	SDI	U	
1	GLUCOSE II	GOD-POD II	Biosystems	310	186.96	8.32	15.56	196 mg/dL	0.58	1.77	
2	UREA II	Urease UV / GLDH II	Biosystems	224	21.96	16.54	3.63	27 mg/dL	1.39	0.49	
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Biosystems	236	6.66	12.14	0.81	7 mg/dL	0.42	0.11	
4	T.BILIRUBIN II	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	63	2.21	15.31	0.34	2.6 mg/dL	1.15	0.09	
5	T-PROTEIN II	Biuret - Colorimetric II	Biosystems	278	5.78	9.57	0.55	5.9 g/dL	0.22	0.07	
6	ALBUMIN II	BCG - Colorimetric II	Biosystems	268	3.38	9.52	0.32	3.2 g/dL	-0.56	0.04	
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Biosystems	260	6.94	11.17	0.78	6.8 mg/dL	-0.18	0.10	
8	CHOLESTROL II	CHOD-PAP II	Biosystems	289	121.21	9.83	11.91	126 mg/dL	0.40	1.40	
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Biosystems	268	132.35	11.45	15.15	148 mg/dL	1.03	1.85	
SDI Range		Interpretation									
Within -1.00 to +1.00		Excellent.									
Within ±1.01 to ±2.00		Good.									
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.									
Beyond ±3.0		Unacceptable performance. Action Signal.									

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

