

Date: 10/10/2023

CALIBRATION CERTIFICATE

This is to certify that **BTS350** Photometer bearing **S. No: 8017511176** installed at **Jayam Clinical Laboratory-Tirukovilur**. has been calibrated on **10/10/2023**

Calibration Values:

LIGHT INTENSITY TEST

LED	TINT	MAIN	REFERENCE
340	39	961812	242453
405	16	971696	739533
505	19	958187	736153
535	19	942163	555641
560	31	982613	777498
600	21	966583	565760
635	16	978971	610629
670	21	963324	706232

PERISTALTIC PUMP ADJUSTMENT TEST

VALUE	CURRENT	NEW
SAMPLE VOLUME	48612	48612
POSITIONING	2100	2100
PUMP DELAY	2	2

Next calibration is due on **09/10/2024**

For **BIOSYSTEMS DIAGNOSTICS PVT. LTD.,**

KRUNAL
GORDHANBHAI
MADHU

Digitally signed by KRUNAL
GORDHANBHAI MADHU
DN: cn=KRUNAL GORDHANBHAI
MADHU, c=IN, o=PERSONAL,
email=krunalmadhu@biosystems.in
Date: 2023.10.10 17:51:06 +05'30'

AUTHORISED SIGNATORY