

**CALIBRATION CERTIFICATE**

**Issued To:** Wellcare Clinical Lab  
SCO 5, Old Ambala Road, Dhakoli, Opposite of  
Civil Hospital Zirakpur  
Chandigarh India 140603

Certificate No: FL/C/OT/DM/03112023-  
C002  
J.O.No: FL/C/OT/DM/03112023-002  
S.R.F.: FL/SRF/20231103/S/AD/15  
Issued Date: 10/11/2023  
Recommended Date for Next Calibration: 03/11/2024



**Details of Instrument** : Optical Microscope  
Make/Model: Ebra / Classic 2020  
ID/S.No: WL/012/Arba Microscope / RH/L/MS/01  
Condition : Working Satisfactory

**Environmental Conditions** : Temperature: (20 ± 2) °C  
Humidity: (50 ± 15) % RH

**Location of Calibration** : On Site, Biochemistry

**Standard (s) used & Uncertainty** : Glass Scale  
Make/ Model/S.No : Mitutoyo / /B1673964

**Traceability of Standard (s)** : TANSON INSTRUMENT  
(Cert. No. TI/E/GS/003/23) Valid upto 10/05/2025

**Support Instrument Used** : Digital Thermo-Hygrometer  
Make/ Model/S.No : FLUKE / 1620 A/C24104

**Methodology Used** : As per Calibration Procedure(s) No. : FL/C/M/SOP-19

**Date of receipt** : 03/11/2023

**Date(s) of Calibration** : 03/11/2023

**Authorized Signatory**  
C.S. Joshi, Director

Page 1 of 2

**NOTE:** The laboratory accepts responsibility for content of this certificate. The results reported in this certificate pertain only to the device calibrated at the time of and under the stated conditions of measurement identified as above. This certificate shall not be reproduced except in full, without written approval of the laboratory. This certificate is intended for guidance and not for legal purpose of advertisement. Any complaint about this certificate should be communicated in writing within 7 days of its issue. Total liability of FARELABS Pvt. Ltd. will be limited to the invoiced amount only. This recipient of this report is requested to confirm the authenticity and genuineness through Email to Fare Labs. In case of any complaint/feedback regarding the calibration of the device, please send an email at [feedback@farelabs.com](mailto:feedback@farelabs.com) or call at +91 9599221227.

**Digitally Signed**

**Issued To:** Wellcare Clinical Lab  
 SCO 5, Old Ambala Road, Dhakoli, Opposite of  
 Civil Hospital Zirakpur  
 Chandigarh India 140603

Certificate No: FL/C/OT/DM/03112023-  
 C002  
 J.O.No: FL/C/OT/DM/03112023-002  
 S.R.F.: FL/SRF/20231103/S/AD/15  
 Issued Date: 10/11/2023  
 Recommended Date for Next Calibration: 03/11/2024



**Measurement Results (At X-Axis )**

S. No.	Nominal Value of DUC	Average Reading On Standard	Error In Unit
	mm	mm	mm
1	0.1	0.10	0.00
2	0.2	0.20	0.00
3	0.4	0.40	0.00
4	0.6	0.60	0.00
5	0.8	0.80	0.00
6	1.0	1.00	0.00

**Measurement Results ( At Y- Axis )**

S. No.	Nominal Value of DUC	Average Reading On Standard	Error In Unit
	mm	mm	mm
1	0.1	0.10	0.00
2	0.2	0.20	0.00
3	0.4	0.40	0.00
4	0.6	0.60	0.00
5	0.8	0.80	0.00
6	1.0	0.99	0.01

All measurement results are traceable to the International system of units (SI), ensured by comparison, directly or indirectly with National or International standards through unbroken chain of calibration. The reported uncertainty is at the coverage factor k=2 which corresponds to a coverage probability of approximately 95 % for a normal distribution.

Uncertainty of Measurement :  $\pm 24.00 \mu m k = 2$  correspondes to confidence level of 95 %.

**DUC\*:** Device Under Calibration

The next date of calibration is recommended as per the client's request.



**Authorized Signatory**  
 C.S.Joshi, Director

**NOTE:** The laboratory accepts responsibility for content of this certificate. The results reported in this certificate pertain only to the device calibrated at the time of and under the stated conditions of measurement identified as above. this certificate shall not be reproduced except in full, without written approval of the laboratory. this certificate is intended of only guidance and not for legal purpose of for advertisement. Any complaint about this certificate should be communicated in writing within 7 days of its issue. Total liability of FARELABS Pvt. Ltd. will be limited to the invoiced amount only. This recipient of this report is requested to confirm the authenticity and genuineness through Email to Fare Labs. In case of any complaint/feedback regarding the calibration of the device, please send an email at [feedback@farelabs.com](mailto:feedback@farelabs.com) or call at +91 9599221227.

----- End of Certificate -----

**Digitally Signed**