

CALIBRATION REPORT

Refrigerators, Laboratory(Lab refrigerator)



| | | | |
|------------|---------------------|-----------------|---------------|
| Report No: | CAL-AR-2022-12-0238 | Date:13-01-2023 | Page 01 of 02 |
|------------|---------------------|-----------------|---------------|

| Customer Details | | | |
|------------------|-------------------|----------|---------------------|
| Health Facility | Rohing CH | | |
| State | Anunachal Pradesh | District | Lower Dibang Valley |

| Medical Equipment Details | | | |
|---------------------------|--|---------------------|------------|
| ME Number | ARZ1LDVDH0060133 | Manufacturer | Videocon |
| ME Name | Refrigerators, Laboratory(Lab refrigerator) | Model | R-134A |
| Department | Laboratory-2 Aids-Opd Block | Serial No | NA |
| Calibrated at | On Site | Condition of ME | Good |
| Date of Calibration | 07-01-2023 | Due for Calibration | 06-01-2024 |

| Environmental Conditions | | | |
|--------------------------|------|------------------|-----|
| Temperature in °C | 20°C | Humidity in % Rh | 58% |

| Details of Master Equipment Used for Calibration | | | | | | |
|--|--------------|---------------------|--------------|---------|------------|----------------------|
| No | Equipment ID | Description | Manufacturer | Model | Serial No. | Calibration Validity |
| 1 | TRI-TE-0189 | Digital Thermometer | METRAVI | DTM-900 | 271926/1 | 15-12-2023 |

| Performance Results | | | | | | |
|---------------------|-------------|----------|------|----------------|------------------|---------|
| No | Parameter | Set Temp | Unit | Measured Value | Acceptable Range | Remarks |
| 1 | Temperature | 2°C | °C | 2.1 | 1.5 to 2.5°C | Pass |
| 2 | Temperature | 4°C | °C | 4.2 | 3.5 to 4.5°C | Pass |
| 3 | Temperature | 6°C | °C | 6.3 | 5.5 to 6.5°C | Pass |

| Conditions of Issue of report |
|-------------------------------|
|-------------------------------|

CALIBRATION REPORT

Refrigerators, Laboratory(Lab refrigerator)



1. The Standards used are traceable to National standards.
2. This report should not be reproduced except in full without written approval from Trimed Solutions India Private Limited.

Calibration Status

Passed

Failed

Calibrated By

Ganesh Sonar

Authorised Signatory

Rahul Kumar Kashyap