

CALIBRATION REPORT



Centrifuge Table top

Report No:	CAL-AR-2022-12-0239	Date:18-12-2022	Page 01 of 02
------------	---------------------	-----------------	---------------

Customer Details			
Health Facility	Roing DH		
State	Arunachal Pradesh	District	Lower Dibang Valley

Medical Equipment Details			
ME Number	ARZ1LOVDH0060134	Manufacturer	Systemex
ME Name	Centrifuges, Tabletop (Centrifuge)	Model	NA
Department	Laboratory-2 Aids-Opd Block	Serial No	NA
Calibrated at	On Site	Condition of ME	Good
Date of Calibration	09-01-2023	Due for Calibration	08-12-2023

Environmental Conditions			
Temperature in °C	20°C	Humidity in % Rh	58%

Details of Master Equipment Used for Calibration						
No	Equipment ID	Description	Manufacturer	Model	Serial No.	Calibration Validity
1	TRI-TE-0015	Digital Tachometer	AMBROBE	TACH-10	1606002 1	15-11-2023

Performance Results						
No	Parameters	Unit	Set Value	Measured Value	Acceptable Range	Remarks
1	Speed	rpm	1500	1530	1425 to 1575	Pass
2	Speed	rpm	2500	2545	2375 to 2625	Pass

CALIBRATION REPORT

Centrifuge Table top



3	Speed	rpm	3500	3600	3325 to 3675	Pass
---	-------	-----	------	------	--------------	------

Conditions of issue of report

1. The Standards used are traceable to National standards.
2. This report should not be reproduced except in full without written approval.

Calibration Status

Passed

Failed

Calibrated By

Ganesh Sonar

Authorized Signatory

Rahul Kumar Kasyp