

VIEW LAB MONTHLY SUMMARY

Lab Name SHRI HANUMAN BALAJI CHARITABLE DIAGNOSTIC CENTER
 Lab No 11428
 Month October
 Year 2023
 Constituent Group Chemistry I

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 18/10/2023

Date of Report Published : 04/11/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	292	334.78	3.17	10.60	333 mg/dL	-0.17	1.24
2	UREA	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	308	44.43	4.48	1.99	43 mg/dL	-0.72	0.23
3	CREATININE	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	313	7.19	4.33	0.31	7.1 mg/dL	-0.29	0.04
4	T.BILIRUBIN	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	305	1.21	10.76	0.13	1.2 mg/dL	-0.08	0.01
5	T-PROTEIN	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	320	4.68	3.63	0.17	4.5 g/dL	-1.06	0.02
6	ALBUMIN	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	304	2.72	5.22	0.14	2.7 g/dL	-0.14	0.02
7	CALCIUM	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	308	11.57	3.83	0.44	11.6 mg/dL	0.07	0.05
8	URIC ACID	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	315	9.80	3.27	0.32	9.7 mg/dL	-0.31	0.04
9	CHOLESTEROL	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	287	91.97	5.21	4.79	90 mg/dL	-0.41	0.57
10	TRIGLYCERIDE	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	283	89.18	4.91	4.38	87 mg/dL	-0.50	0.52
11	HDL	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	286	20.08	6.28	1.26	20 mg/dL	-0.06	0.15
12	SODIUM	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	248	136.69	2.23	3.05	134 mmol/L	-0.88	0.39
13	POTASSIUM	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	246	5.97	2.58	0.15	5.3 mmol/L	-1.10	0.02
14	CHLORIDE	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	210	101.69	2.58	2.62	99 mmol/L	-1.02	0.36
15	AST	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	324	117.35	6.61	7.76	116 U/L	-0.17	0.86
16	ALT	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	323	142.19	9.52	13.54	150 U/L	0.58	1.51
17	ALP	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	312	324.62	8.48	27.52	338 U/L	0.49	3.12

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

Aug 5
 2023

VIEW LAB MONTHLY SUMMARY

Lab Name **SHRI HANUMAN BALAJI CHARITABLE DIAGNOSTIC CENTER**
 Lab No **11428**
 Month **September**
 Year **2023**
 Constituent Group **Chemistry I**

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Click on the analyte to view Graphical Data

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 17/09/2023

Date of Report Published : 05/10/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No. of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	315	192.95	3.60	6.95	193 mg/dL	0.01	0.78
2	UREA	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	319	112.91	3.54	3.99	115 mg/dL	0.52	0.45
3	CREATININE	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	323	7.61	4.26	0.32	7.3 mg/dL	-0.96	0.04
4	T.BILIRUBIN	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	316	1.91	8.26	0.16	2.1 mg/dL	1.20	0.02
5	T-PROTEIN	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	321	5.24	3.93	0.21	5.3 g/dL	0.29	0.02
6	ALBUMIN	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	321	3.03	5.15	0.16	3.2 g/dL	1.09	0.02
7	CALCIUM	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	306	10.09	3.17	0.32	10.1 mg/dL	0.03	0.04
8	URIC ACID	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	315	6.53	3.33	0.22	6.4 mg/dL	-0.60	0.02
9	CHOLESTEROL	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	298	114.62	4.93	5.66	111 mg/dL	-0.64	0.66
10	TRIGLYCERIDE	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	299	134.13	5.72	7.68	130 mg/dL	-0.54	0.89
11	HDL	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	292	22.16	6.53	1.45	21 mg/dL	-0.80	0.17
12	SODIUM	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	256	132.10	2.37	3.13	130 mmol/L	-0.67	0.39
13	POTASSIUM	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	256	4.93	2.72	0.13	5 mmol/L	0.52	0.02
14	CHLORIDE	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	218	102.11	2.62	2.67	100 mmol/L	-0.79	0.36
15	AST	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	326	114.50	6.27	7.18	113 U/L	-0.21	0.79
16	ALT	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	326	59.52	8.18	4.87	60 U/L	0.10	0.54
17	ALP	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	317	123.29	7.59	9.36	121 U/L	-0.24	1.05

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

(Am) (8/10/23)