

### Centrifuge Calibration Verification Record Sheet

Location of the ICTC/PPTCT : ICTC, CHC, Chandragiri, Chittoor District

Equipment ID: ICTC/CHANDRAGIRI/CF/01

Range : up to 3500RPM

Procedure using centrifuge: In house verification of centrifuge using certified standard calibrated by Tachometer

Date of Calibration: 24.08.2023

Due date of Calibration: 23.08.2024

Certified Tachometer Id: SRL/HIVTL/Tachometer

Certified Tachometer calibration date : 30.03.23

Verification done by: G.Archana, Technical Officer,  
Dept. of Microbiology, S.V. Medical College, Tirupati

Sno	Standard reading (RPM)	UUC Reading (RPM)	Error (RPM)
1	499.6	500	0.4
2	999.3	1000	0.7
3	1997.8	2000	2.2
4	2997.4	3000	2.6
5	3497.1	3500	2.9

Accept/ Reject: Accepted since average difference is within +- 100RPM

Final Interpretation: Centrifuge verification is satisfactory

Action taken: NA

Remarks: NA

*Archana*  
Signature

Technical Officer  
APSAES (SRL)  
Department of Microbiology  
S.V. Medical College/SVRRGGH  
TIRUPATI - 517 507