



# S.M. ENTERPRISES

(Calibration and Testing of Electrical, Electronic & Precision Instruments)

27<sup>th</sup>  
Years of Service

Office/Cal. Lab : Plot No. 12, "Ganesh Niwas", Sanjeevaiah Nagar, Co-operative Society, Besides Tadbund Hanuman Temple, Sikh Village, Secunderabad - 500 009, T.S, INDIA (T: (040) 27840697, 48541008  
Cell: 9848020798, 9948291008, 9948292008, 9848293008, 9948294008 Email: msudheer.98@yahoo.com, msudheer.98@smentp.org, SME.003@YAHOO.IN, url: www.smentp.org

## CALIBRATION CERTIFICATE

CALIBRATION ON 25/05/2023		CALIBRATION DUE 24/05/2024	
1. Calibration Report Ref. No. & Date :		SME/NP/SVMC/TIM/03/2023;Dt:25/05/2023	
2. Requested by : (Name and address of the organization)		To, Prof & HOD (State reference Lab) Department Of Microbiology SV Medical College Tirupathi AP	
3. Instrument Name:		Digital Timer	
4. Instrument Make/Model:		Acetecq / DT-1045	
5. Instrument ID. No:		SVMC/TIM/03	
6. Range :		0 to 99 min	
7. Accuracy :		+/- 1 % FSD	
8. Environmental Condition		Temp/RH : 28 +/- 2 Deg C; R.H: 65 +/- 5%	
9. Test parameter & condition :		Time V/S Time (Comparison Method)	
10. Standard Instruments used for Calibration:		Time Interval Meter	
11. Traceability Certificate Ref. No.		C-220622-5-1 ,Date: 29/06/2022 , Due Dt: 29/06/2023	
<b>TEST RESULTS</b>			
SI No	Standard Reading Of Time	Observed Reading Of Time	Error Value(+/-)
	Hr:min:sec:ms	min:sec	Hr:min:sec:ms
1	00:01:00:49	01:00	00:00:00:49
2	00:10:00:38	10:00	00:00:00:38
3	00:30:00:54	30:00	00:00:00:54
4	01:50:00:38	90:00	00:00:00:38
PLEASE NOTE: THE INSTRUMENTS ERROR FOUND TO BE WITHIN THE SPECIFIED ACCURACY LIMIT AND THIS CERTIFICATE VALID FOR PERIOD OF ONE YEAR FROM THE DATE OF ISSUE.			

GSTIN : 36AFVPS4924F1ZW Service Tax No. : AFVPS4924FST001 PAN No. : AFVPS4924F  
NABL APPROVED LAB (CC-2471) / Approved Laboratory of ISO 9001-2015  
TEST EQUIPT. / STANDARDS USED ARE TRACEABLE TO NATIONAL STANDARDS