



ISO 9001 2015
RQ91/3144

S.M. ENTERPRISES

(Calibration and Testing of Electrical, Electronic & Precision Instruments)

27th
Years of Service

Office/Cal. Lab : Plot No. 12 "Ganesh Niwas", Sanjeeviah Nagar Co-operative Society, Besides Tarbund Hanuman Temple, Sakthi Village, Secunderabad - 500 009, T.S, INDIA. z: (040) 27840697, 48541008
Cell: 9848020798, 9948291008, 9948292008, 9848293008, 9948294008 Email: msudheer_98@yahoo.com, msudheer_98@smentp.org, SME_003@YAHOO.IN, url: www.smentp.org

CALIBRATION CERTIFICATE

CALIBRATION ON 25/05/2023	CALIBRATION DUE 24/05/2024
1. Calibration Report Ref. No. & Date :	SME/NP/SVMC/TIM/015/2023;Dt:25/05/2023
2. Requested by : (Name and address of the organization)	To, Prof & HOD (State reference Lab) Department Of Microbiology SV Medical College Tirupathi AP
3. Instrument Name:	Digital Timer
4. Instrument Make/Model:	Aceteq / DT-1045
5. Instrument ID. No:	SVMC/TIM/015
6. Range :	0 to 99 min
7. Accuracy :	+/- 1 % FSD
8. Environmental Condition	Temp/RH : 28 +/- 2 Deg C; R.H: 65 +/- 5%
9. Test parameter & condition :	Time V/S Time (Comparison Method)
10. Standard Instruments used for Calibration:	Time Interval Meter
11. Traceability Certificate Ref. No.	C-220622-5-1 , Done Date: 29/06/2022 , Due Dt: 29/06/2023

TEST RESULTS

Sl No	Standard Reading Of Time Hr:min:sec:ms	Observed Reading Of Time min:sec	Error Value(+/-) Hr:min:sec:ms
1	00:01:00:19	01:00	00:00:00:19
2	00:10:00:45	10:00	00:00:00:45
3	00:30:00:37	30:00	00:00:00:37
4	01:50:00:34	90:00	00:00:00:34

PLEASE NOTE: THE INSTRUMENTS ERROR FOUND TO BE WITHIN THE SPECIFIED ACCURACY LIMIT AND THIS CERTIFICATE VALID FOR PERIOD OF ONE YEAR FROM THE DATE OF ISSUE.

CALIBRATED BY

S. M. Enterprises

QA INCHARGE



GSTIN : 36AFVPS4924F1ZW Service Tax No. : AFVPS4924FST001 PAN No. : AFVPS4924F
NABL APPROVED LAB (CC-2471) / Approved Laboratory of ISO 9001-2015
TEST EQUIPT. / STANDARDS USED ARE TRACEABLE TO NATIONAL STANDARDS