

CENTRIFUGE Calibration / Verification Record Sheet			
Location: ICTC-CHC-PUTTUR		Range : Up to 3500RPM	
Equipment ID: ICTC/PUTTUR/CF/01			
Calibration Procedure Reference: In-house verification of centrifuge using certified standard calibrated by Tachometer			
Date of Calibration : 19.11.2023		Due date of Calibration:18.11.2024	
Certified Tachpmeter ID : SRL/HIVTL/Tachometer		Certified Tachometer Calibration date : 30.03.2023	
Verification Done by: G.ARCHANA , TECHNICAL OFFICER, SRL, DEPARTMENT OF MICROBIOLOGY , S.V.MEDICAL COLLEGE , TIRUPATI			
S.NO	STANDARD READING(RPM)	UUC READING (RPM)	ERROR (RPM)
1.	499.7	500	0.3
2.	999.2	1000	0.8
3.	1998.7	2000	1.3
4	2997.5	3000	2.5
5.	3496.8	3500	3.2
Accept/ Reject: Accepted since average difference is within +/- 100RPM			
Final Interpretation: Centrifuge is Satisfactory			
Action Taken: - NA			
Remark: - NA			

*Archana*  
Signature

**Technical officer,  
SRL, Department of Microbiology  
S.V.Medical College , Tirupati**