

VIEW LAB MONTHLY SUMMARY

| Lab Name | PROMILA DIAGNOSTIC CENTRE | | Details About Robust Analysis | | | | | | | |
|---|---------------------------------|---|---|-----------------------|--------|--------------|-------|-------------|-------|------|
| Lab No | 5290 | | Detail About Monthly Summary | | | | | | | |
| Month | September | | Detail about SDI | | | | | | | |
| Year | 2023 | | | | | | | | | |
| Constituent Group | Chemistry II | | | | | | | | | |
| Click on the analyte to view Graphical Data | | | All Analyser Result | Print | | | | | | |
| Date of Result Entered : 13/09/2023 | | | | | | | | | | |
| Date of Report Published : 05/10/2023 | | | | | | | | | | |
| Sl.No | Analyte | Method / Principle Name | Analyzer Name | No of Participants | DV | Participants | | Your Value | SDI | U |
| | | | | | | CV | SD | | | |
| 1 | GLUCOSE II | GOD-POD II | Mindray | 199 | 343.84 | 9.48 | 32.59 | 346.7 mg/dL | 0.09 | 4.62 |
| 2 | UREA II | Urease UV / GLDH II | Mindray | 157 | 37.80 | 8.61 | 3.25 | 35.6 mg/dL | -0.68 | 0.52 |
| 3 | CREATININE II | Enzymatic Colorimetric II | Mindray | 57 | 1.12 | 12.58 | 0.14 | 0.98 mg/dL | -0.99 | 0.04 |
| 4 | T.BILIRUBIN II | Diazonium Salt (Colorimetric) / Jendrassik II | Mindray | 194 | 5.32 | 12.06 | 0.64 | 5.88 mg/dL | 0.87 | 0.09 |
| 5 | T-PROTEIN II | Biuret - Colorimetric II | Mindray | 188 | 5.46 | 9.20 | 0.50 | 3.98 g/dL | -2.95 | 0.07 |
| 6 | ALBUMIN II | BCG - Colorimetric II | Mindray | 188 | 3.31 | 7.59 | 0.25 | 2.94 g/dL | -1.47 | 0.04 |
| 7 | URIC ACID II | Enzymatic / Uricase Colorimetric II | Mindray | 181 | 9.99 | 8.15 | 0.82 | 10.4 mg/dL | 0.50 | 0.12 |
| 8 | CHOLESTROL II | CHOD-PAP II | Mindray | 208 | 115.97 | 8.47 | 9.83 | 112.5 mg/dL | -0.35 | 1.36 |
| 9 | TRIGLYCERIDE II | GPO-PAP / Enzymatic Colorimetric / End Point II | Mindray | 182 | 208.24 | 7.62 | 15.87 | 201.6 mg/dL | -0.42 | 2.35 |
| SDI Range | | Interpretation | | | | | | | | |
| Within -1.00 to +1.00 | | Excellent. | | | | | | | | |
| Within ±1.01 to ±2.00 | | Good. | | | | | | | | |
| Within ±2.01 to ±2.99 | | Accept with caution. Warning Signal. | | | | | | | | |
| Beyond ±3.0 | | Unacceptable performance. Action Signal. | | | | | | | | |

Links

| | | | |
|--------------------------------|---------------------------------|------------------------------------|--|
| Modify Address | Modify Methods | Submit Lab Results | View Lab Monthly Summary |
| Yearly Summary | Change Password | Home | Sign-out |