



**PROFICIENCY TESTING REPORT**  
**ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME**  
 NABL accredited program as per ISO/IEC 17043:2010 standard  
 Organized By Department of Hematology, AIIMS, New Delhi-110029



*Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens*

EQAP CODE No. : 5820

Distribution No.: 160-0

Month/Year: July/2023

Instrument ID: PDC01

Name & Contact No. of PT Co-ordinator: Dr. Manoranjan Mahapatra ( Prof. & Head), Hematology, AIIMS, Delhi,  
 Tel: 9013085730 , E-Mail : accuracy2000@gmail.com

Date of issue &amp; status of the report: 19-09-2023[Final].

### CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 <sup>3</sup> /µl	1	3.43	3.41	6.84	6.85	0.043	-0.01	0.02	0.08	0.006	-0.81
RBC x10 <sup>6</sup> /µl	1	4.93	4.81	9.74	9.51	0.016	0.74	0.12	0.05	0.004	1.35
Hb g/dl	1	12.95	12.8	25.75	25.3	0.036	0.55	0.15	0.1	0.011	0.34
HCT%	1	43.2	42	85.2	80.7	0.256	0.84	1.2	0.4	0.045	1.35
MCV-fl	1	87.6	87.3	174.9	169.3	0.451	0.57	0.3	0.3	0.030	0.00
MCH-Pg	1	26.6	26.3	52.9	53	0.104	-0.05	0.3	0.2	0.018	0.45
MCHC-g/dl	1	30.5	30	60.5	63.1	0.231	-0.57	0.5	0.3	0.027	0.54
Plt. x10 <sup>3</sup> /µl	1	156	150	306	318	2.317	-0.23	6	7	0.578	-0.15
Retic %	2	0.7	0.5	1.2	1.59	0.054	-0.28	0.2	0.2	0.014	0.00

### P.S . Assesment

YOUR REPORT			CONSENSUS REPORT		
DLC%	3	Nrbcs= , Poly=42 L=03, E=5, Mono/Promono=1 , B1=3 P.M.=5, Mye=70, Meta=12, Other=	Poly: 37 - 52, Myelo: 15 - 27, Meta: 9- 17, Promyelo: 2-8, Lympho: 2- 5, Blast: 1-4, Eosino: 1-3, Mono: 1-2, nRBC/ Baso: 0-5		
RBC Morphology	3	NCNC+,ANISO+POIK+,HYPO+	Predominantly: Normocytic/Normochromic; Moderate: Anisocytosis, hypochromic, Mild: Poikilocytosis		
Diagnosis	3	CML	Chronic Myeloid Leukemia (Chronic Phase)		

**COMBINED DATA VALUES OF TOTAL PARTICIPANTS**

Test parameters	S.No.	Total participants covered in the current dist. 160--O	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
<b>WBC x10<sup>3</sup>/µl</b>	1	200	199	86.43	92.96	3.52	1.01	10.05	6.03
<b>RBC x10<sup>6</sup>/µl</b>	1	200	200	87	86.5	6	5.5	7	8
<b>Hb g/dl</b>	1	200	200	86.5	87.5	7	4.5	6.5	8
<b>HCT%</b>	1	200	199	92.96	89.95	2.51	3.02	4.53	7.03
<b>MCV-fl</b>	1	200	199	92.46	87.44	6.53	4.02	1.01	8.54
<b>MCH-Pg</b>	1	200	199	83.42	92.46	9.05	1.01	7.53	6.53
<b>MCHC-g/dl</b>	1	200	199	90.95	85.43	5.03	5.03	4.02	9.54
<b>Plt. x10<sup>3</sup>/µl</b>	1	200	199	88.94	92.96	6.53	3.52	4.53	3.52
<b>ReticCount%</b>	2	200	152	84.87	92.76	7.24	9.87	7.89	-2.63
<b>PS Assessment</b>	3	200	163	Satisfactory :98.5%, Borderline Sat. :0%, Unsatisfactory :1.50%					

**\*Comments:**

**1). Among Lab (EQA) : Results acceptable.**

**2). Within Lab (IQA) : Precision acceptable.**

**Note-1: EQA** (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

**IQA** ( Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

**Note-2:** Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

**Note-3:** Z score 0 to  $\pm 2$ : Acceptable, Z score  $\pm 2$  to  $\pm 3$  :Warning Signal, Z score  $> \pm 3$  : Unacceptable [As per ISO/IEC 13528:2015 standard]

**Note-4:** Z score value between "0 to  $\pm 2$ " are texted in green colour. Z score value between " $\pm 2$  to  $\pm 3$ " are texted in orange colour. Z score value  $> \pm 3$  are texted in red colour.

**Note-5:** Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value (0.3\*SDPA). To pass the stability test, average difference in measurement values of first and last day sample ( $\bar{x}-\bar{y}$ ) should be smaller than the check value (0.3\*SDPA).

**Note-6:** ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

**Note-7:** Participants are free to use methods/analyzer of their own choice.

**Note-8:** Proficiency testing (PT ) samples are sent quarterly to each participant.

**Note-9:** All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website [www.ishtmaiimseqap.com](http://www.ishtmaiimseqap.com).

**Note 10:** Reports are kept confidential.

Report authorized by,



Dr. Manoranjan Mahapatra ( Prof. & Head)

PT Co-ordinator: ISHTM-AIIMS-EQAP

Department of Hematology, AIIMS, New Delhi

-----End Of Report-----