



## CERTIFICATE OF CALIBRATION

COMMITTED TO THE CUSTOMER SINCE - 1996

No. 301/A, 9th Main Road, 3rd Cross, Rajiv Gandhi Nagar, J.B. Kaval, Nandhini Layout Post, Bangalore - 560 096.  
Ph : 080-23377266, Mob : 9986586789 / 9632221171 / 9964308118 | Email : info@viplgroup.com Web : www.viplgroup.com



NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2017 With vide Certificate No: CC-2473

**1 Name and Address of the Customer** : M/s. LEELAVATI MULTI SPECIALITY HOSPITAL  
7.2nd Main Road Chamaraj Pet  
Bangalore 560 018.

**2 Customer Reference**  
2.1 ULR No : CC247323100011867F  
2.2 SRF No : 8937  
2.3 Certificate No. : VI/22-23/3937-01  
2.4 Format No. : VI-FRM-ME-073  
2.5 Dc No & Dc Date : 55,113 & 27-03-2023  
2.6 Receipt Date : 28-03-2023  
2.7 Date Of Issue : 30-03-2023

**3 Details Of Device Under Calibration(DUC).**  
3.1 Nomenclature : Micro Pipette  
3.2 Make /Model. : Vertex  
3.3 Range/Resolution : 5 - 50µl  
3.4 SL No. : 220801233  
3.5 DUC Condition : Satisfactory  
3.6 Calibration Procedure No. : SOP-16-71 Based On ISO 8655-6:2002 E  
3.7 No.of Pages : 2  
3.8 Calibration Date : 30-03-2023  
3.9 Calibration Due : 29-03-2024  
3.10 Calibration done at : VI Volumetric Lab  
3.11 Discipline : Mechanical (Mass & volume)

**4 Environmental Condition**  
Temperature 20.5 - 20.2 °C Humidity 49.7 - 49.8 %Rh

**5 Standards Used for calibration**

Sl. No.	Nomenclature	Make & Model	SL No& ID No	Range	Certificate. No.	Traceability to	Validity
1	Digital Weighing Balance	Radwag / MYA5.4Y	544953 / VI/ME/DWB/02	up to 5g	VI/22-23/INT-ME-121	VI Bangalore	15-09-2023
2	Digital Weighing Balance	Radwag/XA 82/220.4Y	551565 / VI/ME/DWB-03	up to 220g	VI/22-23/INT-ME-118	VI Bangalore	15-09-2023
3	E2 Class Weight	LCGC & --	2020136 & VI-ME-E2C-001	1mg to 200g	VI/22-23/INT-ME-186-01	VI Bangalore	15-09-2023

**6 Note:**

- 6.1. The Calibration Certificate relates only to the above DUC
- 6.2. Publication or reproduction of this Certificate in any form other than by complete set of the whole report & in the language, written, is not permitted without the written consent of VI Lab..
- 6.3. Corrections/erasing, invalidate the Calibration Certificate.
- 6.4. Calibration of the DUC are traceable to National standards/International Standards
- 6.5. Any error in this Certificate should be brought to our knowledge within 30 days from the date of this Cert.
- 6.6. Results Reported are valid at the time of and under the stated conditions of measurements.
- 6.7. The usage of NABL symbol is as per NABL guidelines given in NABL-133.

Calibrated By

Checked By

Authorized By

P Ramesh Kumar  
(Calibration Engineer)

P. Santhosh Kumar  
(Lab- In-Charge)







COMMITTED TO THE  
CUSTOMER SINCE - 1996

# Vaidyanatheshwara INSTRUMENTS

## CERTIFICATE OF CALIBRATION



CC-2473



No. 301/A, 9th Main Road, 3rd Cross, Rajiv Gandhi Nagar, J.B. Kaval, Nandhini Layout Post, Bangalore - 560 096.  
Ph : 080-23377266, Mob : 9986586789 / 9632221171 / 9964308118 | Email : info@viplgroup.com Web : www.viplgroup.com

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2017 With vide Certificate No: CC-2473

Page 1 of 2

**1 Name and Address of the Customer** : M/s. LEELAVATHI MULTI SPECIALITY HOSPITAL  
7.2nd Main Road Kamalanagar Chamraj Pet  
Bangaluru - 560 016.

**2 Customer Reference**

2.1 ULR No : CC247323100011278F  
2.2 SRF No : 8815  
2.3 Certificate No. : VI/22-23/8815-02  
2.4 Format No. : VI-FRM-ME-073  
2.5 Dc No & Dc Date : -- & 23-03-2023  
2.6 Receipt Date : 24-03-2023  
2.7 Date Of Issue : 24-03-2023

**3 Details Of Device Under Calibration(DUC).**

3.1 Nomenclature : Micro Pipette  
3.2 Make /Model. : Vertex  
3.3 Range/Resolution : 100 - 1000µl /10µl  
3.4 SL No. : 041712840  
3.5 DUC Condition : Satisfactory  
3.6 Calibration Procedure No. : SOP-16-71 Based On ISO 8655-6:2002 E  
3.7 No.of Pages : 2  
3.8 Calibration Date : 24-03-2023  
3.9 Calibration Due : 23-03-2024  
3.10 Calibration done at : VI Volumetric Lab  
3.11 Discipline : Mechanical (Mass & volume)

**4 Environmental Condition**

Temperature 20.5 - 20.2 °C Humidity 49.7 - 49.8 %Rh

**5 Standards Used for calibration**

Sl. No.	Nomenclature	Make & Model	SL No& ID No	Range	Certificate. No.	Traceability to	Validity
1	Digital Weighing Balance	Radwag/XA 82/220.4Y	551565 / VI/ME/DWB-03	up to 220g	VI/22-23/INT-ME-118	VI Bangalore	15-09-2023
2	E2 Class Weight	LCGC & --	2020136 & VI-ME-E2C-001	1mg to 200g	VI/22-23/INT-ME-186-01	VI Bangalore	15-09-2023

**6 Note:**

- The Calibration Certificate relates only to the above DUC
- Publication or reproduction of this Certificate in any form other than by complete set of the whole report & in the language, written, is not permitted without the written consent of VI Lab..
- Corrections/erasing, invalidate the Calibration Certificate.
- Calibration of the DUC are traceable to National standards/International Standards
- Any error in this Certificate should be brought to our knowledge within 30 days from the date of this Cert.
- Results Reported are valid at the time of and under the stated conditions of measurements.
- The usage of NABL symbol is as per NABL guidelines given in NABL-133.

Calibrated By

Checked By

P Ranjith kumar  
(Calibration Engineer)

P. Santhosh Kumar  
(Lab- In-Charge)





## CERTIFICATE OF CALIBRATION

No. 301/A, 9th Main Road, 3rd Cross, Rajiv Gandhi Nagar, J.B. Kaval, Nandhini Layout Post, Bangalore - 560 096.  
Ph : 080-23377266, Mob : 9986586789 / 9632221171 / 9964308118 | Email : info@viplgroup.com Web : www.viplgroup.com

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2017 With vide Certificate No: CC-2473

**1 Name and Address of the Customer** : M/s. LEELAVATHI MULTI SPECIALITY HOSPITAL  
7.2nd Main Road Kamalanagar Chamaraj Pet  
Bengaluru - 560 016.

**2 Customer Reference**

- 2.1 ULR No : CC247323100011277F
- 2.2 SRF No : 8815
- 2.3 Certificate No. : VI/22-23/8815-01
- 2.4 Format No. : VI-FRM-ME-073
- 2.5 Dc No & Dc Date : -- & 23-03-2023
- 2.6 Receipt Date : 24-03-2023
- 2.7 Date Of Issue : 24-03-2023

**3 Details Of Device Under Calibration(DUC).**

- 3.1 Nomenclature : Micro Pipette
- 3.2 Make /Model. : Vertex
- 3.3 Range/Resolution : 5 - 50µl /0.1µl
- 3.4 SL No. : 041713009
- 3.5 DUC Condition : Satisfactory
- 3.6 Calibration Procedure No. : SOP-16-71 Based On ISO 8655-6:2002 E
- 3.7 No.of Pages : 2
- 3.8 Calibration Date : 24-03-2023
- 3.9 Calibration Due : 23-03-2024
- 3.10 Calibration done at : VI Volumetric Lab
- 3.11 Discipline : Mechanical (Mass & volume)

**4 Environmental Condition**

Temperature 20.5 - 20.2 °C Humidity 49.7 - 49.8 %Rh

**5 Standards Used for calibration**

Sl. No.	Nomenclature	Make & Model	SL No& ID No	Range	Certificate. No.	Traceability to	Validity
1	Digital Weighing Balance	Radwag / MYA5.4Y	544953 / VI/ME/DWB/02	up to 5g	VI/22-23/INT-ME-121	VI Bangalore	15-09-2023
2	Digital Weighing Balance	Radwag/XA 82/220.4Y	551565 / VI/ME/DWB-03	up to 220g	VI/22-23/INT-ME-118	VI Bangalore	15-09-2023
3	E2 Class Weight	LCGC & --	2020136 & VI-ME-E2C-001	1mg to 200g	VI/22-23/INT-ME-186-01	VI Bangalore	15-09-2023

**6 Note:**

- 6.1. The Calibration Certificate relates only to the above DUC
- 6.2. Publication or reproduction of this Certificate in any form other than by complete set of the whole report & in the language, written, is not permitted without the written consent of VI Lab..
- 6.3. Corrections/erasing, invalidate the Calibration Certificate.
- 6.4. Calibration of the DUC are traceable to National standards/International Standards
- 6.5. Any error in this Certificate should be brought to our knowledge within 30 days from the date of this Cert.
- 6.6. Results Reported are valid at the time of and under the stated conditions of measurements.
- 6.7. The usage of NABL symbol is as per NABL guidelines given in NABL-133.

Calibrated By

Checked By

Authorized By

P Rangish Kumar  
(Calibration Engineer)

P. Santosh Kumar  
(Lab- In-Charge)

