

20-02-2023

CALIBRATION CERTIFICATE


Name of the customer	Amar Hospital Patiala
Location	Patiala
Make	Beckman Coulter
Product Name	DXH500
Serial No	AZ110611

This is to certify that the cell counter installed at the above mentioned customer site has been calibrated with (Lot no-492316000 dt- 17-FEB.-2023)

S.No	Procedure	Parameter	Remarks
1	Startup	WBC/RBC/HGB/MCV/RDW/PLT/MPV	PASS
2	Reproducibility	WBC/RBC/HGg/MCV/RDW/PLT/MPV	PASS
3	QCTri Level	WBC/RBC/HGB/MCV/RDW/PLT/MPV	PASS
4	Calibration	WBC/RBC/HGB/MCV/ PLT/MPV	PASS

Date of Calibration: 17 FEB.2023	Due Date of Calibration: 16 FEB.2024
----------------------------------	--------------------------------------

For Able Diagnostic Systems Pvt. Ltd


Arzoo Gupta
Asst. Business Manager

