

## CALIBRATION CERTIFICATE

CERTIFICATE NO: SBS/CL/23/17346		MEDICAL DEVICES	Page No:1 of 1
Issue Date	04-11-2023		
SRF No & Date	SRF/23/01037-0002 & 03-11-2023		
Receipt Date	03-11-2023		
Calibration Date	03-11-2023		
Calibration Due	02-11-2024		

**Customer Name & Address**  
 GOVERNMENT PRIMARY HEALTH CENTRE,  
 KANNANUR-621206,TRICHY DISTRICT.

### Details of Device Under Calibration (DUC)

Description : ELECTRICAL SAFETY(BIO CHEMISTRY ANALYZER)	Make & Model : ROBONIK & PRIETEST TOUCH
Range : MULTI	Sr. No : ATCD2770818RBK
Resolution : MULTI	Identification No : NA
DUC Condition : SATISFACTORY	Location : LABORATORY

### Environmental Conditions & Calibration Procedure Details

Environmental Details	Temperature: 25.6°C	Relative Humidity	52% RH
Calibration Procedure No	SBS/CP/MD/29	Calibration done at	ONSITE

### Reference Standards Details

S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	Rigel Medical & 44L-1059	M-230809-16-4	10-08-2024

### ELECTRICAL SAFETY


#### RESULTS

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resisitance >20MΩ	Measured values in MΩ 91	Uncertainty in % ( ± ) 13.92
2	Earth Leakage <5000µAfor B,BF,CF	Measured values in µA 181	Uncertainty in % ( ± ) 9.7
3	Enclosure Leakage <500µAfor B,BF,CF	Measured values in µA 237	Uncertainty in % ( ± ) 8.5

#### REMARKS


1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
2. The user should determine the suitability of the instrument for its intended use.
3. The recalibration interval should be determined on the user requirement.
4. The results stated in this certificate relate only to the item calibrated.
5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

  
 (Calibration Engineer)  
 M.DINESH



Authorised Signatory

  
 Technical Manager  
 (C.SHANMUGARAJ)

Chief Executive

## SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poombagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,  
 E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.