

# CALIBRATION CERTIFICATE

Certificate No: SBS/CL/23/17348

Page No: 1 of 1

Customer Name & Address

GOVERNMENT PRIMARY HEALTH CENTRE,  
KANNANUR-621206, TRICHY DISTRICT.

SRF No.	SRF/23/01037-0004
SRF Date	03-11-2023
Date of Receipt	03-11-2023
Date of Calibration	03-11-2023
Due Date for Calibration	02-11-2024
Issue Date	04-11-2023

**Details of Unit Under Calibration**

Description	Micro Pipette	Make	NA
Range	10µl FIXED	Model	NA
Resolution		Material	PVC
Serial Number	NA	Operating Range	10µl FIXED
ID Number	NA	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

**Environmental Condition**

**Calibration Method Used**

Temperature (°C)	23.9	Humidity (%RH)	55	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08

**Standard Used**

Sl. No.	Description	ID.No. / Sl. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSPL22/12/2115-01	A&D & GH-252	National Standards	09-12-2023

Z Factor: 1.00319

**Result of Calibration in µl**

Sl. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)
1	10	9.94	9.97	9.94	9.96	9.93	9.95	-0.05	0.02	0.47
		9.98	9.95	9.92	9.97	9.96				

**Remarks**

1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
2. The user should determine the suitability of the instrument for its intended use.
3. The recalibration interval should be determined on the user requirement.
4. The results stated in this certificate relate only to the item calibrated.
5. Equipment used for Calibration were calibrated & traceable to National & International Standards
6. The indicated uncertainties are expanded uncertainty estimated at a confidence level of approximately 95% for a coverage factor k=2.00.
7. Calibration Liquid Used: Distilled or Deionized water as per the Code specified in ISO 3696.

Calibrated By,

*M. Racul*  
(Calibration Engineer)  
M.RACUL



Authorised by:

*C. Sivabalan*  
(Quality Management Chief Executive)  
C.SIVABALAN

## SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087.  
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.