



CALIBRATION CERTIFICATE

Page No. 1 of 1

Certificate No: SBS/CI/23/17347	
Customer Name & Address GOVERNMENT PRIMARY HEALTH CENTRE, KANNANUR-621206, TRICHY DISTRICT.	SRF No. SBS/23/01037-0003
	SRF Date 03-11-2023
	Date of Receipt 03-11-2023
	Date of Calibration 03-11-2023
	Due Date for Calibration 02-11-2024
	Issue Date 04-11-2023

Details of Unit Under Calibration			
Description	Micro Pipette	Make	THERMO SCIENTIFIC
Range	100-1000µl	Model	FINNPIPETTE F3
Resolution	1µl	Material	PVC
Serial Number	NA	Operating Range	100-1000µl
ID Number	NA	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition				Calibration Method Used	
Temperature (°C)	23.9	Humidity (%RH)	55	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08

Standard Used						
Sl. No.	Description	ID.No. / Sl. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSPL22/12/2115-01	A&D & GH-252	National Standards	09-12-2023

Z Factor: 1.00319

Result of Calibration in µl											
Sl. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (t)	
1	100	99.85	99.87	99.85	99.84	99.89	99.87	-0.13	0.02	0.47	
		99.89	99.88	99.90	99.87	99.90					
2	500	499.90	499.88	499.90	499.90	499.87	499.88	-0.12	0.02	0.47	
		499.89	499.88	499.85	499.85	499.90					
3	1000	999.89	999.88	999.89	999.87	999.90	999.89	-0.11	0.01	0.47	
		999.88	999.90	999.91	999.86	999.89					

- Remarks
- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
 - The user should determine the suitability of the instrument for its intended use.
 - The recalibration interval should be determined on the user requirement.
 - The results stated in this certificate relate only to the item calibrated.
 - Equipment used for Calibration were calibrated & traceable to National & International Standards
 - The indicated uncertainties are expanded uncertainty for a confidence level of approximately 95% for a coverage factor k=2.00.
 - Calibration Liquid Used: Distilled or Deionized water conforming to ISO 3696.

Calibrated By,

 (Calibration Engineer)
 MERAGUL



Authorised by:

 (Quality Manager/Chief Executive)
 C.SIVABALAN

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor; Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,
 E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.