



REMI SALES & ENGINEERING LIMITED

8/1, 3rd Cross, Shamanna Garden,
Adugodi Post, Behind MICO Back Gate,
Bengaluru - 560 030.
Office : 080- 22224406 / 23
Godown : 080- 22222060
Email : Instrumentbang@remigroup.com
motorsbang@remigroup.com
fanbang@remigroup.com
Website : www.remilabworld.com
www.remigroup.com
CIN : U31100MH1980PLC022314
GST No. : 29AAACR0421M1ZF

CALIBRATION CERTIFICATE

CUSTOMER

LUPIN HELATH CARE LIMITED
KORAMANGALA
BANGALORE

Certificate No: 91/2022-23
Date : 08.03.2023

DESCRIPTION OF ITEM: LABORATORY CENTRIFUGE

MODEL NO: R- 8C

SL.NO. : ZIIN-37371

DETAILS OF EQUIPMENT USED FOR CALIBRATION:

SR.	DESCRIPTION	CERTIFIED BY	CERTIFICATE NO.	VALIDITY
1	DIGITAL TACHOMETER	BELTRONICS LAB	SRF NO. 5236-03	30.06.2023


OBSERVATIONS:

SR.	INSTRUMENT SPEED DISPLAYED RPM	SPEED READING ON DIGITAL TACHO METER RPM	DIFFERENCE IN ACTUAL & TACHO METER RPM	REMARKS
1	1500	1510	10	O.K
2	4500	4520	20	O.K

NEXT CALIBRATION DUE ON: 07.03.2024

- REMARKS: 1) CALIBRATION TRACEABLE TO NATIONAL STANDARDS.
2) OVERALL DEVIATIONS IS WITHIN SPECIFIED ACCURACY.
3) RESULTS ARE RELATED ONLY TO THE ITEM CALIBRATED.
4) MAXIMUM TOLERANCE PERMISSIBLE IS +/- 100RPM.

CALIBRATED BY: SUBHAN



Corporate Office : Remi House Plot No. 11, Cama Road, Estate, Walbhat Road, Goregaon, (E) Mumbai - 400 063. India. Tel : +91-22-40589888
Branches : Ahmedabad, Bengaluru Chennai Goa, Hyderabad, Indore, Jaipur, Kolkata, Kochi, Kanpur, Mumbai, Nagpur, New Delhi, Vijayawada