

# AMS Healthcare

( Erstwhile : Ace Medical Solutions (India) Pvt. Ltd.)

An ISO 9001-2015 Certified Company

No. 770/1, Jothi Complex 2nd Floor, Alagar Kovil Main Road, Surveyor Colony,  
Madurai - 625 007, Tamilnadu, India Mobile : +91 94434 74610



GSTIN: 33AOSPA9153Q1ZA • DL No. TN-12-20B-00398, TN-12-21B-00398

## CALIBRATION CERTIFICATE

Customer Name & Address:	M/S. SUN Clinical Laboratory, #424, Mangammal Salai, K.T.C. Nagar, Tirunelveli.	Certificate No:	AMS/LAB/289/5186
		Issuing Date:	30/10/2023
		Page No:	1 of 1

### Details of Device under Calibration (DUC):

DUC:	Micropipette	Id. No:	SC/LAB/VMP/002
Make:	Dragon Lab	Department:	Laboratory
Model:	---	Calibrated On:	28/10/2023
Sl. No:	YE19BAL0612907	Calibration Due Suggested:	27/10/2024
Type:	Variable	DUC Condition on Receipt:	OK
Range & Least Count:	100 to 1000µl		

### Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Electronic Weighing Scale	EWC/SF400C	PC/01/23/ATL/0512-02	31/01/2024
2	Thermo Hygrometer	Beetech/088	PC/01/23/ATL/0210-02	31/01/2024

Room Temperature: 25°C±2

Room Humidity: 30-60%

(At the time of Calibration)

### Test Results Volume µl:

SL. No.	DUC Reading (µl)	Standard Reading (µl)		Correction (µl)
		Trial 1	Trial 2	
1	100	100.0	100.0	+00.0
2	200	200.0	200.0	+00.0
3	500	500.0	500.0	+00.0
4	800	800.0	800.0	+00.0
5	1000	1000.0	1000.0	+00.0

Allowable Tolerance: User Defined

### Note:

1. Calibration Points were selected as per customer request.
2. Each reading was taken by Repeatability Test(i.e.) each reading in a row is a mean of 10 Readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

### Calibrated By:

(Calibration Engineer)



### Approved by:

(Technical Manager)