

AMS Healthcare

(Erstwhile : Ace Medical Solutions (India) Pvt. Ltd.)

An ISO 9001-2015 Certified Company

No. 770/1, Jothi Complex 2nd Floor, Alagar Kovil Main Road, Surveyor Colony,
Madurai - 625 007, Tamilnadu, India Mobile : +91 94434 74610



GSTIN. 33AOSPA9153Q1ZA • DL No. TN-12-20B -00398, TN-12-21B-00398

CALIBRATION CERTIFICATE

Customer Name & Address:	M/S. SUN Clinical Laboratory, #424, Mangammal Salai, K.T.C. Nagar, Tirunelveli.	Certificate No:	AMS/LAB/289/5185
		Issuing Date:	30/10/2023
		Page No:	1 of 1

Details of Device under Calibration (DUC):

DUC:	Micropipette	Id. No:	SC/LAB/VMP/001
Make:	---	Department:	Laboratory
Model:	---	Calibrated On:	28/10/2023
Sl. No:	JR05009	Calibration Due	27/10/2024
Type:	Variable	Suggested:	
Range & Least Count:	20 to 200µl	DUC Condition on Receipt:	OK

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Electronic Weighing Scale	EWC/SF400C	PC/01/23/ATL/0512-02	31/01/2024
2	Thermo Hygrometer	Beetech/088	PC/01/23/ATL/0210-02	31/01/2024

Room Temperature: 25°C±2

Room Humidity: 30-60%

(At the time of Calibration)

Test Results Volume µl:

SL. No.	DUC Reading (µl)	Standard Reading (µl)		Correction (µl)
		Trial 1	Trial 2	
1	20	20.0	20.0	+00.0
2	50	50.0	50.0	+00.0
3	100	100.0	100.0	+00.0
4	200	200.0	200.0	+00.0

Allowable Tolerance: User Defined

Note:

1. Calibration Points were selected as per customer request.
2. Each reading was taken by Repeatability Test(i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:

(Calibration Engineer)



Approved by:

(Technical Manager)